

SOLW 351

6/10/22

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.	Calib. Date:	06/10/2022
Calibration File No.:	02098	Calib. No.:	00054
Certification File No.:	02099	Cert. Date:	06/10/2022
Linearity File No.:	02100	Lin. Date:	06/10/2022
Solution File No.:	02101	Soln. Date:	06/10/2022
Sequential File No.:	02101	File Date:	06/10/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0040
Solution Control Lot:	21480	Expires:	12/28/2023
		Bottle No.:	0393

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:57D		
Control 1 EC	0.102%	13:57D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:58D		
Control 2 EC	0.099%	13:59D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:00D		
Control 3 EC	0.100%	14:00D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DD x KP2-444 (DL)

Changed By:

Last Name: LUTZ

First Name: DENNIS

MI: J

Signature: Tpr I Lutz 7045

Badge No.: 7045

Date: 06/10/2022