

SOLN 442

6/10/22

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
 Location: TOMS RIVER P.D.
 Calibration File No.: 03377 Calib. Date: 06/10/2022 Calib. No.: 00045
 Certification File No.: 03378 Cert. Date: 06/10/2022 Cert. No.: 00032
 Linearity File No.: 03379 Lin. Date: 06/10/2022 Lin. No.: 00032
 Solution File No.: 03380 Soln. Date: 06/10/2022 Soln. No.: 00442
 Sequential File No.: 03380 File Date: 06/10/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386
 Control Solution %: 0.100% Expires: 12/28/2023
 Solution Control Lot: 21480 Bottle No.: 0380

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:30D		
Control 1 EC	0.104%	11:30D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.103%	11:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31D		
Control 2 EC	0.101%	11:32D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.103%	11:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:33D		
Control 3 EC	0.101%	11:33D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.103%	11:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJP2-025 (DL)

Changed By:
 Last Name: LUTZ First Name: DENNIS MI: J
 Signature: Tpr I Dent 7045 Badge No.: 7045
 Date: 06/10/2022