

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03377	Calib. Date: 06/10/2022
Certification File No.:	03378	Calib. No.: 00045
Linearity File No.:	03379	Cert. Date: 06/10/2022
Solution File No.:	03392	Cert. No.: 00032
Sequential File No.:	03392	Lin. Date: 06/10/2022
		Lin. No.: 00032
		Soln. Date: 06/24/2022
		Soln. No.: 00443
		File Date: 06/24/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	21480	Expires: 12/28/2023
		Bottle No.: 0532

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:50D		
Control 1 EC	-.---%	23:51D	21.8°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	23:51D	21.8°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: *P. Perry #365*

Badge No.: 365

Date: 06/25/2022