

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	02098	Calib. Date:	06/10/2022
Certification File No.:	02099	Cert. Date:	06/10/2022
Linearity File No.:	02100	Lin. Date:	06/10/2022
Solution File No.:	02104	Soln. Date:	06/28/2022
Sequential File No.:	02104	File Date:	06/28/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0387
Solution Control Lot:	21480	Expires:	12/28/2023
		Bottle No.:	0385

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:03D		
Control 1 EC	0.103%	15:03D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	15:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:04D		
Control 2 EC	0.102%	15:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	15:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:05D		
Control 3 EC	0.100%	15:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJP2-026 DL

Changed By:

Last Name: LUTZ First Name: DENNIS MI: J

Signature: *Tip I Dent* Badge No.: 7045

Date: 06/28/2022