

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03377	Calib. Date: 06/10/2022
Certification File No.:	03378	Calib. No.: 00045
Linearity File No.:	03379	Cert. Date: 06/10/2022
Solution File No.:	03428	Cert. No.: 00032
Sequential File No.:	03428	Lin. Date: 06/10/2022
		Lin. No.: 00032
		Soln. Date: 07/28/2022
		Soln. No.: 00446
		File Date: 07/28/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	21480	Expires: 12/28/2023
		Bottle No.: 0554

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:38D		
Control 1 EC	0.101%	11:38D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:39D		
Control 2 EC	0.099%	11:40D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:40D		
Control 3 EC	0.100%	11:41D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	11:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NATER

First Name: MARK

MI: S

Signature: \_\_\_\_\_

*PH. Nater* #336

Badge No.: 336

Date: 07/28/2022