

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	02098	Calib. Date: 06/10/2022
Certification File No.:	02099	Calib. No.: 00054
Linearity File No.:	02100	Cert. Date: 06/10/2022
Solution File No.:	02110	Cert. No.: 00041
Sequential File No.:	02110	Lin. Date: 06/10/2022
		Lin. No.: 00041
		Soln. Date: 08/04/2022
		Soln. No.: 00353
		File Date: 08/04/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0387
Solution Control Lot:	21480	Expires: 12/28/2023
		Bottle No.: 0433

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:48D		
Control 1 EC	0.101%	08:49D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	08:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:50D		
Control 2 EC	0.100%	08:50D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	08:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:51D		
Control 3 EC	0.100%	08:52D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	08:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:52D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:  #350

Badge No.: 350

Date: 08/04/2022