

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03377	Calib. Date: 06/10/2022
Certification File No.:	03378	Calib. No.: 00045
Linearity File No.:	03379	Cert. Date: 06/10/2022
Solution File No.:	03445	Cert. No.: 00032
Sequential File No.:	03445	Lin. Date: 06/10/2022
		Lin. No.: 00032
		Soln. Date: 08/14/2022
		Soln. No.: 00447
		File Date: 08/14/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	21480	Expires: 12/28/2023
		Bottle No.: 0435

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:24D		
Control 1 EC	0.101%	09:25D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:26D		
Control 2 EC	0.099%	09:26D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:27D		
Control 3 EC	0.101%	09:28D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: FITZGERALD

First Name: RYAN

MI: P

Signature: _____

 #399

Badge No.: 1399

Date: 08/14/2022