

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	02098	Calib. Date: 06/10/2022
Certification File No.:	02099	Calib. No.: 00054
Linearity File No.:	02100	Cert. Date: 06/10/2022
Solution File No.:	02113	Cert. No.: 00041
Sequential File No.:	02113	Lin. Date: 06/10/2022
		Lin. No.: 00041
		Soln. Date: 08/25/2022
		Soln. No.: 00354
		File Date: 08/25/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0387
Solution Control Lot:	21480	Expires: 12/28/2023
		Bottle No.: 0394

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:53D		
Control 1 EC	0.103%	13:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	13:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:54D		
Control 2 EC	0.100%	13:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	13:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:56D		
Control 3 EC	0.101%	13:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	13:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: NATER

First Name: MARK

MI: S

Signature: _____



Badge No.: 336

Date: 08/25/2022