

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.		
Calibration File No.:	03377	Calib. Date:	06/10/2022
Certification File No.:	03378	Cert. No.:	00045
Linearity File No.:	03379	Cert. Date:	06/10/2022
Solution File No.:	03455	Lin. Date:	06/10/2022
Sequential File No.:	03455	Lin. No.:	00032
		Soln. Date:	08/25/2022
		Soln. No.:	00448
		File Date:	08/25/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0386
Solution Control Lot:	21480	Expires:	12/28/2023
		Bottle No.:	0399

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:51D		
Control 1 EC	0.102%	13:52D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	13:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:53D		
Control 2 EC	0.101%	13:53D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	13:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:54D		
Control 3 EC	0.101%	13:55D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.103%	13:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NATER

First Name: MARK

MI: S

Signature: \_\_\_\_\_

Badge No.: 336

Date: 08/25/2022