

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 03377 Calib. Date: 06/10/2022 Calib. No.: 00045
Certification File No.: 03378 Cert. Date: 06/10/2022 Cert. No.: 00032
Linearity File No.: 03379 Lin. Date: 06/10/2022 Lin. No.: 00032
Solution File No.: 03486 Soln. Date: 09/30/2022 Soln. No.: 00450
Sequential File No.: 03486 File Date: 09/30/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386
Control Solution %: 0.100% Expires: 04/04/2024
Solution Control Lot: 22110 Bottle No.: 1114

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:27D		
Control 1 EC	0.102%	09:27D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	09:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:28D		
Control 2 EC	0.100%	09:29D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	09:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:29D		
Control 3 EC	0.101%	09:30D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:31D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:  4350

Badge No.: 350

Date: 09/30/2022