

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.	Calib. Date:	06/10/2022
Calibration File No.:	02098	Calib. No.:	00054
Certification File No.:	02099	Cert. Date:	06/10/2022
Linearity File No.:	02100	Lin. Date:	06/10/2022
Solution File No.:	02122	Soln. Date:	10/27/2022
Sequential File No.:	02122	File Date:	10/27/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0387
Solution Control Lot:	22110	Expires:	04/04/2024
		Bottle No.:	1221

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:29D		
Control 1 EC	0.102%	09:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:31D		
Control 2 EC	0.100%	09:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:32D		
Control 3 EC	0.100%	09:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	09:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature: _____

Adam Koepen #350

Badge No.: 350

Date: 10/27/2022