

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	02123	Calib. Date:	10/28/2022
Certification File No.:	02124	Cert. No.:	00055
Linearity File No.:	02125	Cert. Date:	10/28/2022
Solution File No.:	02126	Lin. No.:	00042
Sequential File No.:	02126	Lin. Date:	10/28/2022
		Soln. No.:	00357
		Soln. Date:	10/28/2022
		File Date:	10/28/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0387
Solution Control Lot:	22110	Expires:	04/04/2024
		Bottle No.:	0243

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:30D		
Control 1 EC	0.102%	13:31D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:31D		
Control 2 EC	0.100%	13:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:33D		
Control 3 EC	0.100%	13:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJP2-026 DC

**Changed By:**

Last Name: LUTZ      First Name: DENNIS      MI: J

Signature: *Tyr I Lutz*      Badge No.: 7045      Date: 10/28/2022