

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARAJ-0075  
Location: TOMS RIVER P.D.  
Calibration File No.: 03507 Calib. Date: 10/28/2022 Calib. No.: 00046  
Certification File No.: 03508 Cert. Date: 10/28/2022 Cert. No.: 00033  
Linearity File No.: 03509 Lin. Date: 10/28/2022 Lin. No.: 00033  
Solution File No.: 03516 Soln. Date: 11/05/2022 Soln. No.: 00453  
Sequential File No.: 03516 File Date: 11/05/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386  
Control Solution %: 0.100% Expires: 04/04/2024  
Solution Control Lot: 22110 Bottle No.: 0233

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:16S		
Control 1 EC	0.099%	12:17S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:17S		
Control 2 EC	0.097%	12:18S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	12:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:19S		
Control 3 EC	0.096%	12:19S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	12:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:20S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: CASSADY

First Name: MICHAEL

MI: P

Signature: 

Badge No.: 367

Date: 11/05/2022