

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 03507
Certification File No.: 03508
Linearity File No.: 03509
Solution File No.: 03529
Sequential File No.: 03529
Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 22110

Serial No.: ARAJ-0075
Calib. Date: 10/28/2022
Cert. Date: 10/28/2022
Lin. Date: 10/28/2022
Soln. Date: 11/20/2022
File Date: 11/20/2022
Calib. No.: 00046
Cert. No.: 00033
Lin. No.: 00033
Soln. No.: 00456
Model No.: CU-34
Serial No.: DDWK S3-0386
Expires: 04/04/2024
Bottle No.: 0129

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:29S		
Control 1 EC	0.099%	08:30S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:30S		
Control 2 EC	0.096%	08:31S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:32S		
Control 3 EC	0.097%	08:32S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:33S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FITZGERALD

First Name: RYAN

MI: P

Signature:  #399

Badge No.: 1399

Date: 11/20/2022