Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 03507 Calib. Date: 10/28/2022 Calib. No.: 00046
Certification File No.: 03508 Cert. Date: 10/28/2022 Cert. No.: 00033
Linearity File No.: 03509 Lin. Date: 10/28/2022 Lin. No.: 00033
Solution File No.: 03535 Soln. Date: 11/26/2022 Soln. No.: 00458
Sequential File No.: 03535 File Date: 11/26/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386
Control Solution %: 0.100% Expires: 04/04/2024
Solution Control Lot: 22110 Bottle No.: 1226

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:40S		
Control 1 EC	0.100%	08:40S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:41S		
Control 2 EC	0.095%	08:42S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.095%	08:42S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:42S		
Control 3 EC	0.093%	08:43S	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.093%	08:43S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	08:44S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: KOEPPEN First Name: ADAM MI: M
Signature: #350 Badge No.: 350
Date: 11/26/2022