

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	02123	Calib. Date:	10/28/2022
Certification File No.:	02124	Cert. Date:	10/28/2022
Linearity File No.:	02125	Lin. Date:	10/28/2022
Solution File No.:	02146	Soln. Date:	12/15/2022
Sequential File No.:	02146	File Date:	12/15/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0387
Solution Control Lot:	22110	Expires:	04/04/2024
		Bottle No.:	1102

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:46S		
Control 1 EC	0.100%	08:47S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:48S		
Control 2 EC	0.099%	08:48S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:49S		
Control 3 EC	0.099%	08:50S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:50S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:  H 352

Badge No.: 350

Date: 12/15/2022