

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 03507 **Calib. Date:** 10/28/2022 **Calib. No.:** 00046
Certification File No.: 03508 **Cert. Date:** 10/28/2022 **Cert. No.:** 00033
Linearity File No.: 03509 **Lin. Date:** 10/28/2022 **Lin. No.:** 00033
Solution File No.: 03541 **Soln. Date:** 12/15/2022 **Soln. No.:** 00460
Sequential File No.: 03541 **File Date:** 12/15/2022

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDWK S3-0386
Control Solution %: 0.100% **Expires:** 04/04/2024
Solution Control Lot: 22110 **Bottle No.:** 1014

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:40S		
Control 1 EC	0.100%	08:41S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:41S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:42S		
Control 2 EC	0.098%	08:42S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:42S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:43S		
Control 3 EC	0.096%	08:44S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	08:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:44S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature: 

Badge No.: 350

Date: 12/15/2022