

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 03507 Calib. Date: 10/28/2022 Calib. No.: 00046
Certification File No.: 03508 Cert. Date: 10/28/2022 Cert. No.: 00033
Linearity File No.: 03509 Lin. Date: 10/28/2022 Lin. No.: 00033
Solution File No.: 03544 Soln. Date: 12/19/2022 Soln. No.: 00461
Sequential File No.: 03544 File Date: 12/19/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386
Control Solution %: 0.100% Expires: 04/04/2024
Solution Control Lot: 22110 Bottle No.: 1492

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:55S		
Control 1 EC	0.100%	09:55S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:56S		
Control 2 EC	0.098%	09:57S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:57S		
Control 3 EC	0.099%	09:58S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NATER

First Name: MARK

MI: S

Signature:  #336

Badge No.: 336

Date: 12/19/2022