

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARWF-0394 |
| Location: | TOMS RIVER P.D. | |
| Calibration File No.: | 02123 | Calib. Date: 10/28/2022 |
| Certification File No.: | 02124 | Calib. No.: 00055 |
| Linearity File No.: | 02125 | Cert. Date: 10/28/2022 |
| Solution File No.: | 02155 | Cert. No.: 00042 |
| Sequential File No.: | 02155 | Lin. Date: 10/28/2022 |
| | | Lin. No.: 00042 |
| | | Soln. Date: 12/24/2022 |
| | | Soln. No.: 00360 |
| | | File Date: 12/24/2022 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDWK S3-0387 |
| Solution Control Lot: | 22110 | Expires: 04/04/2024 |
| | | Bottle No.: 1331 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 14:25S | | |
| Control 1 EC | 0.099% | 14:25S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 14:25S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:26S | | |
| Control 2 EC | 0.096% | 14:27S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 14:27S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:28S | | |
| Control 3 EC | 0.098% | 14:28S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 14:28S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:29S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CASSADY

First Name: MICHAEL

MI: P

Signature: _____

PH1 M. Cassidy 367

Badge No.: 367

Date: 12/24/2022