

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03507	Calib. Date: 10/28/2022
Certification File No.:	03508	Calib. No.: 00046
Linearity File No.:	03509	Cert. Date: 10/28/2022
Solution File No.:	03548	Cert. No.: 00033
Sequential File No.:	03548	Lin. Date: 10/28/2022
		Lin. No.: 00033
		Soln. Date: 01/11/2023
		Soln. No.: 00463
		File Date: 01/11/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	22110	Expires: 04/04/2024
		Bottle No.: 1340

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:09S		
Control 1 EC	0.103%	09:10S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:10S		
Control 2 EC	0.100%	09:11S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:12S		
Control 3 EC	0.100%	09:12S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:13S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:  # 350

Badge No.: 350

Date: 01/11/2023