

# Calibrating Unit New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03507	Calib. Date: 10/28/2022
Certification File No.:	03508	Calib. No.: 00046
Linearity File No.:	03509	Cert. Date: 10/28/2022
Solution File No.:	03553	Cert. No.: 00033
Sequential File No.:	03553	Lin. Date: 10/28/2022
		Lin. No.: 00033
		Soln. Date: 01/23/2023
		Soln. No.: 00464
		File Date: 01/23/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	22110	Expires: 04/04/2024
		Bottle No.: 0904

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:48S		
Control 1 EC	0.100%	09:49S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:49S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:49S		
Control 2 EC	0.098%	09:50S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:51S		
Control 3 EC	0.098%	09:51S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:51S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:52S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WESTFALL	First Name: ROBERT	MI: S
Signature: 	Badge No.: 306	Date: 01/23/2023