

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03507	Calib. Date: 10/28/2022
Certification File No.:	03508	Calib. No.: 00046
Linearity File No.:	03509	Cert. Date: 10/28/2022
Solution File No.:	03561	Cert. No.: 00033
Sequential File No.:	03561	Lin. Date: 10/28/2022
		Lin. No.: 00033
		Soln. Date: 01/31/2023
		Soln. No.: 00465
		File Date: 01/31/2023

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWK S3-0386
Control Solution %:	0.100%		Expires: 04/04/2024
Solution Control Lot:	22110		Bottle No.: 1218

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:42S		
Control 1 EC	0.098%	16:43S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:43S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:44S		
Control 2 EC	0.097%	16:44S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:45S		
Control 3 EC	0.097%	16:46S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FITZGERALD

First Name: RYAN

MI: P

Signature: _____

RY #399

Badge No.: 1399

Date: 01/31/2023