

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	02123	Calib. Date: 10/28/2022
Certification File No.:	02124	Calib. No.: 00055
Linearity File No.:	02125	Cert. Date: 10/28/2022
Solution File No.:	02179	Cert. No.: 00042
Sequential File No.:	02179	Lin. Date: 10/28/2022
		Lin. No.: 00042
		Soln. Date: 02/04/2023
		Soln. No.: 00362
		File Date: 02/04/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0387
Solution Control Lot:	22110	Expires: 04/04/2024
		Bottle No.: 0244

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:35S		
Control 1 EC	0.099%	12:35S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:36S		
Control 2 EC	0.099%	12:37S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:38S		
Control 3 EC	0.098%	12:38S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CASSADY

First Name: MICHAEL

MI: P

Signature: *PH M. Cassidy 367*

Badge No.: 367

Date: 02/04/2023