

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03507	Calib. Date: 10/28/2022
Certification File No.:	03508	Calib. No.: 00046
Linearity File No.:	03509	Cert. Date: 10/28/2022
Solution File No.:	03574	Cert. No.: 00033
Sequential File No.:	03574	Lin. Date: 10/28/2022
		Lin. No.: 00033
		Soln. Date: 02/14/2023
		Soln. No.: 00467
		File Date: 02/14/2023

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWK S3-0386
Control Solution %:	0.100%		Expires: 04/04/2024
Solution Control Lot:	22110		Bottle No.: 1228

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:57S		
Control 1 EC	0.100%	12:57S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:58S		
Control 2 EC	0.096%	12:59S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:00S		
Control 3 EC	0.097%	13:00S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____



Badge No.: 306

Date: 02/14/2023