

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0394
Location: TOMS RIVER P.D.
Calibration File No.: 02123 Calib. Date: 10/28/2022 Calib. No.: 00055
Certification File No.: 02124 Cert. Date: 10/28/2022 Cert. No.: 00042
Linearity File No.: 02125 Lin. Date: 10/28/2022 Lin. No.: 00042
Solution File No.: 02187 Soln. Date: 02/25/2023 Soln. No.: 00363
Sequential File No.: 02187 File Date: 02/25/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0387
Control Solution %: 0.100% Expires: 04/04/2024
Solution Control Lot: 22110 Bottle No.: 1346

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:50S		
Control 1 EC	0.102%	16:51S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:52S		
Control 2 EC	0.099%	16:53S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:53S		
Control 3 EC	0.100%	16:54S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:55S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:  # 350

Badge No.: 350

Date: 02/25/2023