

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.	Calib. Date:	10/28/2022
Calibration File No.:	02123	Calib. No.:	00055
Certification File No.:	02124	Cert. Date:	10/28/2022
Linearity File No.:	02125	Lin. Date:	10/28/2022
Solution File No.:	02195	Soln. Date:	03/15/2023
Sequential File No.:	02195	File Date:	03/15/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0387
Solution Control Lot:	22110	Expires:	04/04/2024
		Bottle No.:	0906

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:42S		
Control 1 EC	0.100%	00:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:43S		
Control 2 EC	0.099%	00:44S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	00:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:45S		
Control 3 EC	0.098%	00:45S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:46S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARSICANO

First Name: JOHN

MI: R

Signature: _____

John Marsicano #405

Badge No.: 405

Date: 03/15/2023