

SOLN 365
3-20-23

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	02197	Calib. Date:	03/20/2023
Certification File No.:	02198	Cert. No.:	00056
Linearity File No.:	02199	Lin. Date:	03/20/2023
Solution File No.:	02200	Cert. No.:	00043
Sequential File No.:	02200	Lin. No.:	00043
		Soln. Date:	03/20/2023
		Soln. No.:	00365
		File Date:	03/20/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0040
Solution Control Lot:	22110	Expires:	04/04/2024
		Bottle No.:	1011

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:56S	13:56 D DL	
Control 1 EC	0.102%	12:57S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:57S		
Control 2 EC	0.099%	12:58S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:59S		
Control 3 EC	0.099%	13:00S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:01S	14:01 D DL	

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXKP2-444 (DL)

Changed By:
 Last Name: LUTZ First Name: DENNIS MI: J
 Signature: Ter I. Dent 7045 Badge No.: 7045
 Date: 03/20/2023