

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 03606 **Calib. Date:** 03/20/2023 **Calib. No.:** 00047
Certification File No.: 03607 **Cert. Date:** 03/20/2023 **Cert. No.:** 00034
Linearity File No.: 03608 **Lin. Date:** 03/20/2023 **Lin. No.:** 00034
Solution File No.: 03624 **Soln. Date:** 04/09/2023 **Soln. No.:** 00475
Sequential File No.: 03624 **File Date:** 04/09/2023

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAN-0041
Control Solution %: 0.100% **Expires:** 02/01/2025
Solution Control Lot: 23040 **Bottle No.:** 1269

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:01D		
Control 1 EC	0.101%	23:01D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	23:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:02D		
Control 2 EC	0.097%	23:03D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	23:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:03D		
Control 3 EC	0.097%	23:04D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BARTOSHEK

First Name: CPL

MI: D.

Signature: CPL D. Bartoshek #316 TRPD

Badge No.: 316

Date: 04/09/2023



ORIGINAL