

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 03606 Calib. Date: 03/20/2023 Calib. No.: 00047
Certification File No.: 03607 Cert. Date: 03/20/2023 Cert. No.: 00034
Linearity File No.: 03608 Lin. Date: 03/20/2023 Lin. No.: 00034
Solution File No.: 03627 Soln. Date: 04/14/2023 Soln. No.: 00476
Sequential File No.: 03627 File Date: 04/14/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041
Control Solution %: 0.100% Expires: 02/01/2025
Solution Control Lot: 23040 Bottle No.: 1359

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:03D		
Control 1 EC	0.096%	23:03D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:04D		
Control 2 EC	0.096%	23:05D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:06D		
Control 3 EC	0.097%	23:06D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:07D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 04/14/2023