

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	02197	Calib. Date:	03/20/2023
Certification File No.:	02198	Cert. No.:	00056
Linearity File No.:	02199	Cert. Date:	03/20/2023
Solution File No.:	02207	Lin. Date:	03/20/2023
Sequential File No.:	02207	Soln. Date:	04/15/2023
		File Date:	04/15/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0040
Solution Control Lot:	23040	Expires:	02/01/2025
		Bottle No.:	0898

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:29D		
Control 1 EC	0.101%	02:30D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:31D		
Control 2 EC	0.099%	02:32D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:32D		
Control 3 EC	0.097%	02:33D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:

 # 365

Badge No.: 365

Date: 04/15/2023