

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.		
Calibration File No.:	03606	Calib. Date:	03/20/2023
Certification File No.:	03607	Cert. Date:	03/20/2023
Linearity File No.:	03608	Lin. Date:	03/20/2023
Solution File No.:	03638	Soln. Date:	04/22/2023
Sequential File No.:	03638	File Date:	04/22/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0041
Solution Control Lot:	23040	Expires:	02/01/2025
		Bottle No.:	1423

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:37D		
Control 1 EC	0.100%	22:37D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	22:37D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:38D		
Control 2 EC	0.097%	22:39D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:39D		
Control 3 EC	0.096%	22:40D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:

 #365

Badge No.: 365

Date: 04/22/2023