

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.		
Calibration File No.:	03606	Calib. Date:	03/20/2023
Certification File No.:	03607	Cert. No.:	00047
Linearity File No.:	03608	Lin. Date:	03/20/2023
Solution File No.:	03641	Cert. No.:	00034
Sequential File No.:	03641	Lin. No.:	00034
		Soln. Date:	04/23/2023
		Soln. No.:	00481
		File Date:	04/23/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0041
Solution Control Lot:	23040	Expires:	02/01/2025
		Bottle No.:	1124

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:50D		
Control 1 EC	0.097%	07:50D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:51D		
Control 2 EC	0.096%	07:52D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:52D		
Control 3 EC	0.096%	07:53D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:54D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: 

Badge No.: 365

Date: 04/23/2023