

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03606	Calib. Date: 03/20/2023
Certification File No.:	03607	Calib. No.: 00047
Linearity File No.:	03608	Cert. Date: 03/20/2023
Solution File No.:	03665	Cert. No.: 00034
Sequential File No.:	03665	Lin. Date: 03/20/2023
		Lin. No.: 00034
		Soln. Date: 05/28/2023
		Soln. No.: 00486
		File Date: 05/28/2023

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDAN-0041
Control Solution %:	0.100%		Expires: 02/01/2025
Solution Control Lot:	23040		Bottle No.: 1353

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:06D		
Control 1 EC	0.097%	00:06D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	00:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:07D		
Control 2 EC	0.096%	00:08D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	00:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:09D		
Control 3 EC	0.097%	00:09D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	00:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: 

Badge No.: 365

Date: 05/28/2023