

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03606	Calib. Date: 03/20/2023
Certification File No.:	03607	Calib. No.: 00047
Linearity File No.:	03608	Cert. Date: 03/20/2023
Solution File No.:	03676	Cert. No.: 00034
Sequential File No.:	03676	Lin. Date: 03/20/2023
		Lin. No.: 00034
		Soln. Date: 06/04/2023
		Soln. No.: 00489
		File Date: 06/04/2023

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDAN-0041
Control Solution %:	0.100%		Expires: 02/01/2025
Solution Control Lot:	23040		Bottle No.: 1017

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:05D		
Control 1 EC	0.096%	01:06D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	01:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:07D		
Control 2 EC	0.095%	01:07D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	01:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:08D		
Control 3 EC	0.095%	01:09D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	01:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:09D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 06/04/2023