

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	02197	Calib. Date: 03/20/2023
Certification File No.:	02198	Calib. No.: 00056
Linearity File No.:	02199	Cert. Date: 03/20/2023
Solution File No.:	02254	Cert. No.: 00043
Sequential File No.:	02254	Lin. Date: 03/20/2023
		Lin. No.: 00043
		Soln. Date: 06/10/2023
		Soln. No.: 00369
		File Date: 06/10/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0040
Solution Control Lot:	23040	Expires: 02/01/2025
		Bottle No.: 1362

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:57D		
Control 1 EC	0.101%	07:57D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:58D		
Control 2 EC	0.097%	07:59D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:00D		
Control 3 EC	0.099%	08:00D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:01D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 06/10/2023