

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03606	Calib. Date: 03/20/2023
Certification File No.:	03607	Calib. No.: 00047
Linearity File No.:	03608	Cert. Date: 03/20/2023
Solution File No.:	03701	Cert. No.: 00034
Sequential File No.:	03701	Lin. Date: 03/20/2023
		Lin. No.: 00034
		Soln. Date: 07/09/2023
		Soln. No.: 00497
		File Date: 07/09/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0041
Solution Control Lot:	23010	Expires: 01/11/2025
		Bottle No.: 1202

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:45D		
Control 1 EC	0.095%	09:46D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	09:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:46D		
Control 2 EC	0.094%	09:47D	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.097%	09:47D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	09:48D		
Control 3 EC	0.094%	09:48D	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.097%	09:48D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	09:49D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FITZGERALD

First Name: RYAN

MI: P

Signature: _____

[Handwritten Signature] #399

Badge No.: 1399

Date: 07/09/2023