

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03606	Calib. Date: 03/20/2023
Certification File No.:	03607	Calib. No.: 00047
Linearity File No.:	03608	Cert. Date: 03/20/2023
Solution File No.:	03703	Cert. No.: 00034
Sequential File No.:	03703	Lin. Date: 03/20/2023
		Lin. No.: 00034
		Soln. Date: 07/10/2023
		Soln. No.: 00499
		File Date: 07/10/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0041
Solution Control Lot:	23010	Expires: 01/11/2025
		Bottle No.: 1355

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:23D		
Control 1 EC	0.096%	10:24D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	10:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:24D		
Control 2 EC	0.094%	10:25D	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.096%	10:25D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	10:26D		
Control 3 EC	0.094%	10:26D	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.097%	10:26D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	10:27D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SCHWARTZ

First Name: STEVE

MI: -

Signature: _____



Badge No.: 1293

Date: 07/10/2023