

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	02287	Calib. Date:	07/26/2023
Certification File No.:	02288	Cert. No.:	00057
Linearity File No.:	02289	Lin. Date:	07/26/2023
Solution File No.:	02290	Cert. No.:	00044
Sequential File No.:	02290	Lin. No.:	00044
		Soln. No.:	00372
		File Date:	07/26/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0040
Solution Control Lot:	23010	Expires:	01/11/2025
		Bottle No.:	1376

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:37D		
Control 1 EC	0.101%	13:38D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:39D		
Control 2 EC	0.099%	13:39D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	13:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:40D		
Control 3 EC	0.099%	13:41D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DPXK92-444 (24)

Changed By:

Last Name: LUTZ First Name: DENNIS MI: J

Signature: *Tr I Lutz* 7045 Badge No.: 7045 Date: 07/26/2023