

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.		
Calibration File No.:	03704	Calib. Date:	07/26/2023
Certification File No.:	03705	Cert. No.:	00048
Linearity File No.:	03706	Cert. Date:	07/26/2023
Solution File No.:	03707	Lin. Date:	07/26/2023
Sequential File No.:	03707	Soln. Date:	07/26/2023
		File Date:	07/26/2023

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDAN-0041
Control Solution %:	0.100%			Expires:	01/11/2025
Solution Control Lot:	23010			Bottle No.:	0846

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:04D		
Control 1 EC	0.103%	11:05D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.103%	11:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:06D		
Control 2 EC	0.101%	11:06D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.103%	11:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:07D		
Control 3 EC	0.101%	11:07D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.103%	11:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:08D		


All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXKP2-436 DC

Changed By:

Last Name: LUTZ First Name: DENNIS MI: J

Signature:  Badge No.: 7045
 Date: 7045 Date: 07/26/2023