

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03704	Calib. Date: 07/26/2023
Certification File No.:	03705	Calib. No.: 00048
Linearity File No.:	03706	Cert. Date: 07/26/2023
Solution File No.:	03714	Cert. No.: 00035
Sequential File No.:	03714	Lin. Date: 07/26/2023
		Lin. No.: 00035
		Soln. Date: 08/12/2023
		Soln. No.: 00501
		File Date: 08/12/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0041
Solution Control Lot:	23010	Expires: 01/11/2025
		Bottle No.: 1272

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:06D		
Control 1 EC	0.100%	04:07D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:08D		
Control 2 EC	0.097%	04:08D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	04:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:09D		
Control 3 EC	0.098%	04:10D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	04:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:10D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 08/12/2023