

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03704	Calib. Date: 07/26/2023
Certification File No.:	03705	Calib. No.: 00048
Linearity File No.:	03706	Cert. Date: 07/26/2023
Solution File No.:	03722	Cert. No.: 00035
Sequential File No.:	03722	Lin. Date: 07/26/2023
		Lin. No.: 00035
		Soln. Date: 08/30/2023
		Soln. No.: 00504
		File Date: 08/30/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0041
Solution Control Lot:	23040	Expires: 02/01/2025
		Bottle No.: 0885

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:24D		
Control 1 EC	0.098%	15:24D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	15:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:25D		
Control 2 EC	0.097%	15:25D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:26D		
Control 3 EC	0.098%	15:27D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: \_\_\_\_\_



Badge No.: 306

Date: 08/30/2023