

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0394
Location: TOMS RIVER P.D.
Calibration File No.: 02287 Calib. Date: 07/26/2023 Calib. No.: 00057
Certification File No.: 02288 Cert. Date: 07/26/2023 Cert. No.: 00044
Linearity File No.: 02289 Lin. Date: 07/26/2023 Lin. No.: 00044
Solution File No.: 02313 Soln. Date: 09/11/2023 Soln. No.: 00374
Sequential File No.: 02313 File Date: 09/11/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0040
Control Solution %: 0.100% Expires: 02/01/2025
Solution Control Lot: 23040 Bottle No.: 1230

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:11D		
Control 1 EC	0.101%	15:12D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:13D		
Control 2 EC	0.099%	15:13D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:14D		
Control 3 EC	0.098%	15:15D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTEALL

First Name: ROBERT

MI: S

Signature: _____



Badge No.: 306

Date: 09/11/2023