

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.		
Calibration File No.:	03704	Calib. Date:	07/26/2023
Certification File No.:	03705	Cert. Date:	07/26/2023
Linearity File No.:	03706	Lin. Date:	07/26/2023
Solution File No.:	03726	Soln. Date:	09/11/2023
Sequential File No.:	03726	File Date:	09/11/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0041
Solution Control Lot:	23040	Expires:	02/01/2025
		Bottle No.:	1228

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:19D		
Control 1 EC	0.101%	15:20D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:20D		
Control 2 EC	0.099%	15:21D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:22D		
Control 3 EC	0.100%	15:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.



Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____

Badge No.: 306

Date: 09/11/2023