

# Calibrating Unit

## New Standard Solution Report

|                         |                       |                         |
|-------------------------|-----------------------|-------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.: ARAJ-0075   |
| Location:               | TOMS RIVER P.D.       |                         |
| Calibration File No.:   | 03704                 | Calib. Date: 07/26/2023 |
| Certification File No.: | 03705                 | Calib. No.: 00048       |
| Linearity File No.:     | 03706                 | Cert. Date: 07/26/2023  |
| Solution File No.:      | 03730                 | Cert. No.: 00035        |
| Sequential File No.:    | 03730                 | Lin. Date: 07/26/2023   |
|                         |                       | Lin. No.: 00035         |
|                         |                       | Soln. Date: 09/28/2023  |
|                         |                       | Soln. No.: 00507        |
|                         |                       | File Date: 09/28/2023   |
| <br>                    |                       |                         |
| Calibrating Unit:       | WET                   | Model No.: CU-34        |
| Control Solution %:     | 0.100%                | Serial No.: DDAN-0041   |
| Solution Control Lot:   | 23040                 | Expires: 02/01/2025     |
|                         |                       | Bottle No.: 1095        |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 01:10D |                |                     |
| Control 1 EC      | 0.097% | 01:10D | 33.9°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.097% | 01:10D | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:11D |                |                     |
| Control 2 EC      | 0.095% | 01:11D | 33.9°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.098% | 01:11D | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:12D |                |                     |
| Control 3 EC      | 0.096% | 01:13D | 33.9°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.098% | 01:13D | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:14D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: \_\_\_\_\_

*Pascal Gambardella #365*

Badge No.: 365

Date: 09/28/2023