

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	02287	Calib. Date:	07/26/2023
Certification File No.:	02288	Cert. Date:	07/26/2023
Linearity File No.:	02289	Lin. Date:	07/26/2023
Solution File No.:	02325	Soln. Date:	09/30/2023
Sequential File No.:	02325	File Date:	09/30/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0040
Solution Control Lot:	23040	Expires:	02/01/2025
		Bottle No.:	1331

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:24D		
Control 1 EC	0.101%	02:24D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	02:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:25D		
Control 2 EC	0.099%	02:26D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	02:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:26D		
Control 3 EC	0.098%	02:27D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:28D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365-

Badge No.: 365

Date: 09/30/2023