

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03704	Calib. Date: 07/26/2023
Certification File No.:	03705	Calib. No.: 00048
Linearity File No.:	03706	Cert. Date: 07/26/2023
Solution File No.:	03739	Cert. No.: 00035
Sequential File No.:	03739	Lin. Date: 07/26/2023
		Lin. No.: 00035
		Soln. Date: 11/05/2023
		Soln. No.: 00509
		File Date: 11/05/2023

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDAN-0041
Control Solution %:	0.100%		Expires: 02/01/2025
Solution Control Lot:	23040		Bottle No.: 1397

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:24S		
Control 1 EC	0.101%	06:25S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	06:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:25S		
Control 2 EC	0.099%	06:26S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.103%	06:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:27S		
Control 3 EC	0.100%	06:27S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	06:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: _____

P. P. Gambardella #365

Badge No.: 365

Date: 11/05/2023