

ALCOTEST CERTIFICATIONS INDEX # 02

02-06-2024

ALCOTEST 9510

SERIAL # ARMJ-0142

- ITEM # 1 9510 PARAMETER REPORT
- ITEM # 2 9510 CERT REPORT – WET ADJUST (PART I)
- ITEM # 3 9510 CERT REPORT – DRY ADJUST (PART II)
- ITEM # 4 9510 CERT REPORT – LINEARITY (PART III)
- ITEM # 5 9510 CYLINDER INSTALLATION REPORT – INLET 1
- ITEM # 6 9510 CYLINDER INSTALLATION REPORT – INLET 2
- ITEM # 7 CERT OF ANALYSIS, CALGAS LOT# 302-402798029
- ITEM # 8 CERT OF ANALYSIS, CALGAS LOT# 302-402571549
- ITEM # 9 CERT OF ACCURACY 9510, SERIAL # ARMJ-0142
- ITEM # 10 CERT OF ANALYSIS SIMULATOR SOLUTION LOT # 22240
- ITEM # 11 CERT OF CALIBRATION, CERT/SO#5-E3K4K-100-1 Revision 0 (5 pgs)
- ITEM # 12 CERT OF CALIBRATION, CERT/SO#5-E3K4K-20-1 Revision 0 (5 pgs)
- ITEM # 13 CERT OF ANALYSIS, CALGAS LOT# 302-402448281
- ITEM # 14 CERT OF ANALYSIS, CALGAS LOT# 302-402488140
- ITEM # 15 CERT OF ANALYSIS, CALGAS LOT# 1474760
- ITEM # 16 CERT OF ANALYSIS, CALGAS LOT# 302-402486003
- ITEM # 17 CERT OF ANALYSIS, CALGAS LOT# 1495468
- ITEM # 18 NJSP CERTIFICATION OFFICER CREDENTIALS

ALCOTEST 9510 PARAMETER REPORT

Equipment

Serial No.: ARMJ-0142
Firmware: 8326739 1.5
WinCE application: 8326738 2.9
Configuration: 8326737 3.10

Date: 11/20/2023
Time: 12:00:41

Parameter

| | | |
|---|--------|-------------|
| min. blow time | 5.0 | s |
| min. breath volume for females of age 60+ | 1.2 | L |
| min. breath volume for all other | 1.5 | L |
| min. blow flow | 4.5 | L/min |
| plateau detection limit | 4 | % |
| plateau detection start conc. | 70 | microgram/L |
| neg. flow detection (part. vacuum) | 10.0 | hPa |
| neg. flow detection sensitivity | 10 | |
| cal. gas abort volume | 0.4 | L |
| result-to-zero limit | 0.0050 | %BAC |
| ambient air check limit | 0.0049 | %BAC |
| interference det. d-criterion limit abs. | 38 | microgram/L |
| interference det. d-criterion limit rel. | 10.0 | % |
| interference det. t-criterion limit abs. | 8 | microgram/L |
| interference det. t-criterion limit rel. | 2.1 | % |
| IR CO2 offset | 10 | microgram/L |
| IR H2O offset | 4 | microgram/L |
| EC H2O offset | 0 | microgram/L |
| Value-based EC aging comp. on/off (1/0) | 0 | |
| Time-based EC aging comp. on/off (1/0) | 1 | |
| Time-based EC aging comp. per month | 0.2 | % |
| Time-based EC aging comp. maximum | 3.0 | % |
| EC fatigue comp. max. sum | 15000 | |
| EC fatigue comp. factor | 50 | |
| EC fatigue comp. minutes | 180 | |
| mouth alc. mark limit | 500 | |
| mouth alc. lower limit | 30 | |
| mouth alc. slope | 6 | |
| mouth alc. zero limit | 50 | |
| mouth alc. max. neg. sum | 6 | |
| mouth alc. max. 2nd derivative | 35 | |

ALCOTEST 9510 CERTIFICATION REPORT - WET ADJUST (PART I)
Toms River Twp

Equipment

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMJ-0142
Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9

Wet Adjust Record

Wet Adjust File No.: 23 Wet Adjust Date: 11/20/2023 Wet Adjust No.: 1
Wet Adjust Time: 13:22:15

Concentration: 0.100 %
Adjusting Unit: X-Cal 2000 Adj. Unit Ser. No.: ARND-0005 Adj. Unit Exp.: 03/15/2024
Solution Lot No.: 22240 Soln. Bottle No.: 879 Adjust Soln. Exp.: 07/05/2024

Preadjust Simulator Temp.: 34.00 degree C
Postadjust Simulator Temp.: 34.01 degree C

Result

Procedure completed successfully.

Coordinator

Last Name: Lutz - First Name: Dennis MI: J Badge No.: 7045

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.



Signed:

Date: 11/20/2023

ID: 5

ALCOTEST 9510 CERTIFICATION REPORT - DRY ADJUST (PART II)
Toms River Twp

Equipment

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMJ-0142
Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9

Dry Adjust Record

Dry Adjust File No.: 24 Dry Adjust Date: 11/20/2023 Dry Adjust No.: 1
Dry Adjust Time: 13:39:56

Concentration: 0.100 %
Dry Gas Lot No.: 302-402448281 Adjust Gas Exp.: 05/19/2025
Barom. Model No.: Mensor CPG2300 Barom. Serial No.: 41001273 Barom. Cert. Exp.: 03/22/2024
Pre-adjust Amb. Pressure: 1023 hPa Post-adjust Amb. Pressure: 1025 hPa

Result

Procedure completed successfully.

Coordinator

Last Name: Lutz - First Name: Dennis MI: J Badge No.: 7045

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

Tpr I Lutz 7045

Signed:

Date: 11/20/2023

ID: 5

ALCOTEST 9510 CERTIFICATION REPORT - LINEARITY (PART III)
Toms River Twp

Equipment

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMJ-0142
Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9

Linearity Record

Linearity File No.: 25 Lin. Date: 11/20/2023 Lin. No.: 1

0.040% Dry Gas Lot No.: 302-402488140 Adjust. Gas Exp.: 07/15/2025
0.080% Dry Gas Lot No.: 1474760 Adjust. Gas Exp.: 04/03/2024
0.160% Dry Gas Lot No.: 302-402488003 Adjust. Gas Exp.: 07/12/2025
0.300% Dry Gas Lot No.: 1495468 Adjust. Gas Exp.: 05/14/2024

Data Summary

| Function | Result %BAC | Time hh:mm:ss | Barometric Pres. [hPa] | Comment(s) or Status Code |
|-----------------------|----------------|------------------|---------------------------|------------------------------|
| Ambient Air Blank | 0.000 | 13:47:36 | | *TEST PASSED* |
| Control .04 Test 1 EC | 0.039 | 13:48:11 | 1025 | *TEST PASSED* |
| Control .04 Test 1 IR | 0.039 | 13:48:11 | 1025 | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 13:49:07 | | *TEST PASSED* |
| Control .04 Test 2 EC | 0.040 | 13:49:19 | 1025 | *TEST PASSED* |
| Control .04 Test 2 IR | 0.039 | 13:49:19 | 1025 | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 13:50:35 | | *TEST PASSED* |
| Control .08 Test 3 EC | 0.078 | 13:51:08 | 1025 | *TEST PASSED* |
| Control .08 Test 3 IR | 0.079 | 13:51:08 | 1025 | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 13:52:08 | | *TEST PASSED* |
| Control .08 Test 4 EC | 0.080 | 13:52:19 | 1025 | *TEST PASSED* |
| Control .08 Test 4 IR | 0.080 | 13:52:19 | 1025 | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 13:53:35 | | *TEST PASSED* |
| Control .16 Test 5 EC | 0.159 | 13:54:08 | 1025 | *TEST PASSED* |
| Control .16 Test 5 IR | 0.160 | 13:54:08 | 1025 | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 13:55:14 | | *TEST PASSED* |
| Control .16 Test 6 EC | 0.161 | 13:55:25 | 1025 | *TEST PASSED* |
| Control .16 Test 6 IR | 0.162 | 13:55:25 | 1025 | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 14:05:36 | | *TEST PASSED* |
| Control .30 Test 7 EC | 0.301 | 14:06:14 | 1025 | *TEST PASSED* |
| Control .30 Test 7 IR | 0.302 | 14:06:14 | 1025 | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 14:07:28 | | *TEST PASSED* |
| Control .30 Test 8 EC | 0.306 | 14:07:43 | 1025 | *TEST PASSED* |
| Control .30 Test 8 IR | 0.306 | 14:07:43 | 1025 | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 14:08:05 | | *TEST PASSED* |

Result

All tests within acceptable tolerance.

Coordinator

Last Name: Lutz - First Name: Dennis MI: J Badge No.: 7045

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

T. Lutz 7045

Signed:

Date: 11/20/2023

ID: 5

ALCOTEST 9510 CYLINDER INSTALLATION REPORT - INLET 1
Toms River Twp
SERIAL NUMBER: ARMJ-0142

Equipment

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMJ-0142
Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9
Cyl1 Install File No.: 96 Cyl1 Install Date: 02/06/2024 Cyl1 Install No.: 2

Control Tests (0.100%)

Installation Inlet: #1 (Upper) Post test active Cyl.: #2 (Lower)
Dry Gas Lot No.: 302-402798029 Dry Gas Lot Exp.: 07/17/2026

Data Summary

| Function | Result %BAC | Time hh:mm:ss | Barometric Pres. [hPa] | Comment(s) or Status Code |
|-------------------|-------------|---------------|------------------------|---------------------------|
| Ambient Air Blank | 0.000 | 12:24:12 | | *TEST PASSED* |
| Control Test 1 | | | 1018 | *TEST PASSED* |
| EC Result | 0.099 | 12:24:58 | | *TEST PASSED* |
| IR Result | 0.103 | 12:24:58 | | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 12:26:16 | | *TEST PASSED* |
| Control Test 2 | | | 1019 | *TEST PASSED* |
| EC Result | 0.101 | 12:26:40 | | *TEST PASSED* |
| IR Result | 0.103 | 12:26:40 | | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 12:28:00 | | *TEST PASSED* |
| Control Test 3 | | | 1018 | *TEST PASSED* |
| EC Result | 0.101 | 12:28:24 | | *TEST PASSED* |
| IR Result | 0.103 | 12:28:24 | | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 12:29:00 | | *TEST PASSED* |

Result

All tests within acceptable tolerance.

Coordinator

Last Name: Lutz - First Name: Dennis MI: J Badge No.: 7045

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

TPF I Dast 7045

Signed:

Date: 02/06/2024

ID: 5

ALCOTEST 9510 CYLINDER INSTALLATION REPORT - INLET 2
Toms River Twp
SERIAL NUMBER: ARMJ-0142

Equipment

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMJ-0142
 Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9
 Cyl2 Install File No.: 27 Cyl2 Install Date: 11/20/2023 Cyl2 Install No.: 1

Control Tests (0.100%)

Installation Inlet: #2 (Lower) Post test active Cyl.: #2 (Lower)
 Dry Gas Lot No.: 302-402571549 Dry Gas Lot Exp.: 10/18/2025

Data Summary

| Function | Result %BAC | Time hh:mm:ss | Barometric Pres. [hPa] | Comment(s) or Status Code |
|-------------------|-------------|---------------|------------------------|---------------------------|
| Ambient Air Blank | 0.000 | 14:20:37 | | *TEST PASSED* |
| Control Test 1 | | | 1025 | *TEST PASSED* |
| EC Result | 0.100 | 14:21:23 | | *TEST PASSED* |
| IR Result | 0.101 | 14:21:23 | | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 14:22:27 | | *TEST PASSED* |
| Control Test 2 | | | 1025 | *TEST PASSED* |
| EC Result | 0.101 | 14:22:50 | | *TEST PASSED* |
| IR Result | 0.101 | 14:22:50 | | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 14:23:55 | | *TEST PASSED* |
| Control Test 3 | | | 1025 | *TEST PASSED* |
| EC Result | 0.102 | 14:24:19 | | *TEST PASSED* |
| IR Result | 0.101 | 14:24:19 | | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 14:24:42 | | *TEST PASSED* |

Result

All tests within acceptable tolerance.

Coordinator

Last Name: Lutz - First Name: Dennis MI: J Badge No.: 7045

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

Teri Lutz 7045

Signed:

Date: 11/20/2023

ID: 5

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Part Number: 4401036

DRAEGER MEDICAL SYSTEMS INC

Sales order: 1122756628

Date: August 10, 2023

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer

ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.

CALGAZ LOT#: 302-402798029

ETHANOL IN NITROGEN

Product Expiration: July 17, 2026

| COMPONENT | PPM | (BrAC) |
|--------------------------|----------|----------|
| ETHANOL | 260.5PPM | (0.100) |
| NITROGEN | BAL | |
| AVERAGE ANALYTICAL VALUE | PPM | (BrAC) |
| ETHANOL | 264.8 | (0.102) |

| REFERENCE STANDARD | CYLINDER | CONCENTRATION PPM |
|-----------------------------|----------|-------------------|
| N.M.I. TRACEABLE STANDARDS* | ND38424 | 260.7 |

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: A679-20190918, D049803-20220329


No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: July 17, 2023

APPROVED BY: _____



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410) 228-6400 Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

DRAEGER INC HOUSTON

Sales order: 1114712164

Date: October 27, 2022

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer
ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.
CALGAZ LOT#: 302-402571549
ETHANOL IN NITROGEN

Product Expiration: October 18, 2025

| COMPONENT | PPM | (BrAC) |
|--------------------------|----------|----------|
| ETHANOL | 260.5PPM | (0.100) |
| NITROGEN | BAL | |
| AVERAGE ANALYTICAL VALUE | PPM | (BrAC) |
| ETHANOL | 262.8 | (0.101) |

| REFERENCE STANDARD | CYLINDER | CONCENTRATION PPM |
|-----------------------------|----------|-------------------|
| N.M.I. TRACEABLE STANDARDS* | ND38434 | 260.4 |

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Calibration test 283192, dated 6th January 2022 or calibration test 292029, 292030 or 292031, dated 26th March 2022 applies

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: ND38434-20211028, A679, A650, ND38462-20211027, ND18363-20211104, ND50144-20201218, ND23368-20211130


No effecting environmental condillons during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: October 18, 2022

APPROVED BY: _____



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 179.85 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410) 228-6400 Fax: (410) 228-4251



Dräger

Alcotest 9510

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 9510 has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest 9510 is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 49 FR 48864, and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your state's specifications.

Certification Date:

Serial Number:

11/1/2023

ARMJ - 0142

Draeger, Inc.

JP MB



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

PHILIP D. MURPHY
Governor

MATTHEW J. PLATKIN
Acting Attorney General

SHEILA Y. OLIVER
Lt. Governor

COLONEL PATRICK J. CALLAHAN
Superintendent

CERTIFICATION OF ANALYSIS
0.100 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.

ANALYSIS DATE: 07/21/2022

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 22240

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1205 to 0.1219 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is July 05, 2024.

As OFS Director for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Michael Kennedy
Michael Kennedy
OFS Director
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 27 day of July, 2022.



KAREN E. STAHL
NOTARY PUBLIC OF NEW JERSEY
Commission # 50110822
My Commission Expires 01/13/2024

NOTARY PUBLIC OF NEW JERSEY
Commission # 60143077
Expires 01/13/2024

"An Internationally Accredited Agency"
New Jersey is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable



Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: 4302958578



Certificate/ISO Number: 5-E3K4K-100-1 Revision 0

Manufacturer: Drager Safety AG & Co. KGaA
Model Number: X-Cal 2000
Description: Breath Alcohol Simulator
Serial Number: ARND-0005
ID: NONE

As-Found: In Tolerance
As-Left: In Tolerance
Issue Date: Mar 15, 2023
Calibration Date: Mar 15, 2023
Due Date: Mar 15, 2024

Calibrated To: Customer Specification
Calibration Procedure: 1-AC103519-1

Transcat Calibration Laboratories have been audited and found in compliance with ISO/IEC 17025:2017. Accredited calibrations performed within the Lab Scope of Accreditation are indicated by the presence of the Accrediting Body Logo and Certificate Number. Any measurements on an accredited calibration not covered by the Lab Scope of Accreditation are listed in the notes section of the certificate. SCC, NRC, CLAS or ANAB do not guarantee the accuracy of an individual calibration by accredited laboratories.

Transcat calibrations, as applicable, are performed in compliance with the requirements of the Transcat Quality Manual QAC-P01-000, the customer Purchase Order and/or Quality Agreement requirements, ISO 9001:2015, ANSI/NCSL Z540.1-1994 (R2002), and ISO 10012:2003, as applicable. When specified contractually, the requirements of ISO TS16949:2009, 10CFR21, 10CFR60 App. B, ASME NQA-1:2012, and ANSI/NCSL Z540.3-2006 (R2013) are also covered.


Complete records of work performed are maintained by Transcat and are available for inspection. Laboratory standards used in the performance of this calibration are listed on this certificate.

Transcat documents the traceability of measurements to the SI units through the National Institute of Standards and Technology (NIST), or the National Research Council of Canada (NRC), or other national measurement institutes (NMI) that are signatories to the CIPM Mutual Recognition Arrangement, or accepted fundamental and/or natural physical constants, or by the use of specified methods, consensus standards or ratio type measurements. Documentation supporting traceability information is available for review upon written request at a Transcat facility. The measured quantity and the measurement uncertainty are required for further dissemination of traceability.

Uncertainties are reported with a coverage factor $k=2$, providing a level of confidence of approximately 95%. All calibrations have been performed using processes having a TUR of 4:1 or better (3:1 for mass calibrations), unless otherwise noted. The Test Uncertainty Ratio (TUR) is calculated in accordance with NCSL International RP-1B. For mass calibrations: Conventional mass referenced to 8.0 g/cm³.

The results in this report relate only to the item calibrated or tested. Recorded calibration data is valid at the time of calibration within the stated uncertainties at the environmental conditions noted. The determination of compliance to the specification is specific to the model/serial no./ID no. referenced above based on the tolerances shown; these tolerances are either the original equipment manufacturers (OEM's) warranted specifications or the client's requested specifications. Any number of factors can cause a unit to drift out of tolerance at any time following its calibration. Limitations on the uses of this instrument are detailed in the OEM's operating instructions. This certificate may not be reproduced except in full, without the written approval of Transcat. Additional information, if applicable may be included on separate report(s).

Customer: DRAEGER INC
 7256 S SAM HOUSTON PKWY W
 STE 100
 HOUSTON, TX 77085
 PO Number: 4302958578



Certificate/SO Number: 5-E3K4K-100-1 Revision 0

As Found/As Left Data

| Description | Setpoints | Accuracy | Low Limit | High Limit | As Found / As Left | Cal Process | | Units | TUR |
|---|-----------|------------|-----------|------------|--------------------|-------------|----------------------------------|-------|---------|
| | | | | | | O | Measurement Uncertainty (k=2; ±) | | |
| Function Checks | | | | | | | | | |
| Bubble Check | | | P | P | P | | | | |
| Seal Check | | | P | P | P | | | | |
| Temperature Source: Accuracy Test | | | | | | | | | |
| Accuracy Test | 34.00°C | ±(0.02 °C) | 33.98 | 34.02 | 34.00 °C | 2.2e-002 | 2.3e-002 | °C | 0.9 : 1 |
| Temperature Source: Stability Test | | | | | | | | | |
| Stability Test | 0.00°C | ±(0.02 °C) | -0.02 | 0.02 | 0.00 °C | 5.0e-003 | 7.6e-003 | °C | 4.0 : 1 |

Field not applicable.

Traceable Standards

| Asset | Manufacturer | Model Number | Description | Cal Date | Due Date | Traceability Number | Use |
|----------|-----------------------|--------------|-------------------|-----------|-----------|---------------------|-------|
| 05H1259 | AccuMac Corporation | AM1760-12-S | Secondary SPRT | 17-Feb-23 | 29-Feb-24 | 15-805H1259-6-1 | AF/AL |
| HP927312 | Hart Scientific/Fluke | 1575 | Super Thermometer | 6-Dec-22 | 30-Jun-24 | 5-&HP927312-8-1 | AF/AL |

The use of the standard is defined as: AF - used for as-found readings, AL - used for as-left readings.

Environmental Data

| Temperature | Relative Humidity | Temp / RH Asset | Lab Area | Lab Description |
|------------------|-------------------|-----------------|----------|-----------------|
| 69.89°F /21.05°C | 34.00% | Dewk15 | G | Temperature |

CALIBRATED
BY **TRANSCAT**

CERTIFICATE OF CALIBRATION

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: 4302958578



Certificate/SO Number: 5-E3K4K-100-1 Revision 0

Decision Rule

When compliance statements are present, they are reported without factoring in the effects of uncertainty and comply with the guidelines as follows: The acceptance zone is defined as: less than or equal to the high limit, and/or greater than or equal to the low limit. The rejection zones are defined as greater than the high limit and/or less than the low limit. Single measurement results in the acceptance zone are identified as in-tolerance. Single measurement results in the rejection zone are identified as out-of-tolerance (OOT). When all measurement results are in the acceptance zone for repeated measurements, for the same characteristic, the test is identified as in-tolerance. For repeated characteristic measurements, a single measurement result in the rejection zone, will cause the test to be identified as out-of-tolerance (OOT). Data rejection for cause, (outliers) is permitted after the "Determining and Verifying Out Of Tolerance (OOT) and/or Op Fail Readings" procedure outlined in this document has been completed and the anomalous reading cannot be repeated, and the anomalous reading does not represent the system under test. Statements of conformity are binary.



Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: 4302968578

Certificate/SO Number: 5-E3K4K-100-1 Revision 0

Legend

| Topic | Description |
|-------------------------------|--|
| Accuracy | UUT specification that establishes expected tolerances and a time limit (calibration interval) over which the instrument is expected to hold these tolerances |
| As Found | Initial measurement results |
| As Left | Measurement results after adjustment and/or repair |
| Blank Data Field | Test is not applicable for the UUT |
| Cal Process Uncertainty (CPU) | The uncertainty of calibration process for the reported measurement result |
| Calibration Date | Indicates the date that the calibration was completed |
| Cover Factor (k) | A measure of uncertainty that defines an interval about the measurement result |
| Due Date | Indicates the end of the calibration cycle as requested by the customer |
| Issue Date | Indicates the date that the calibration has passed the Data Review Process and was signed by an authorized signatory or the date that a revision to the original certificate has been issued |
| Low / High Limits | Establishes UUT acceptable performance limits for the test measurement |
| Measurement Uncertainty | The dispersion of the values attributed to a measured quantity |
| OOA | Out of Acceptance (#) |
| OOT | Out of Tolerance (*) |
| Setpoints | Measurement target values |
| Traceability | Unbroken chain of comparisons relating an instrument's measurements to a known standard(s) |
| Traceability Number | Unique identifier(s) used to document traceability of calibration standards |
| TUR | Test Uncertainty Ratio, ratio of the tolerance or specification of the test measurement in relation to the uncertainty in measurement results |
| UUT | Unit Under test |

CALIBRATED

BY TRANSCAT

CERTIFICATE OF CALIBRATION

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: 4302958578



Certificate/SO Number: 5-E3K4K-100-1 Revision 0


Calibrated At: 16115 Park Row
Houston, TX 77084

Facility Responsible:
16115 Park Row
Houston, TX 77084
800-828-1470


Unit Barcode: 
0900B498610

Date Received: March 13, 2023
Service Level: R9


Calibrated By:

 Jose Martinez
Jose Martinez


Electronically Signed By:

 Jose Martinez
Jose Martinez

Reviewed By:

 Josh Solleau for
Josh Solleau for

Electronically Signed By:

 Josh Solleau for
Josh Solleau for

Jose Martinez
Calibration Technician
Mar 15, 2023
05:25:03 -04:00

Scott D. Caine
Lab Manager
Mar 15, 2023
09:40:28 -04:00

Certificate - Page 5 of 5

Customer Number: 1-659111-00C
OPS-F20-014R10 09/29/21 FP001R9 4/9/2021



Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: 4302958578

Certificate/ISO Number: 5-E3K4K-20-1 Revision 0

Manufacturer: Wika Instr/Mensor Corp/Trend
Model Number: CPG2300
Description: Portable Barometer
Serial Number: 41001273
ID: NONE

As-Found: In Tolerance
As-Left: In Tolerance

Issue Date: Mar 22, 2023
Calibration Date: Mar 22, 2023
Due Date: Mar 22, 2024

Calibrated To: Manufacturer Specification
Calibration Procedure: 1-AC107288-0

Transcat Calibration Laboratories have been audited and found in compliance with ISO/IEC 17025:2017. Accredited calibrations performed within the Lab Scope of Accreditation are indicated by the presence of the Accrediting Body Logo and Certificate Number. Any measurements on an accredited calibration not covered by the Lab Scope of Accreditation are listed in the notes section of the certificate. SCC, NRC, CLAS or ANAB do not guarantee the accuracy of an individual calibration by accredited laboratories.

Transcat calibrations, as applicable, are performed in compliance with the requirements of the Transcat Quality Manual QAC-P01-000, the customer Purchase Order and/or Quality Agreement requirements, ISO 9001:2015, ANSI/NCSL Z540.1-1994 (R2002), and ISO 10012:2003, as applicable. When specified contractually, the requirements of ISO TS16949:2009, 10CFR21, 10CFR50 App. B, ASME NQA-1:2012, and ANSI/NCSL Z540.3-2006 (R2013) are also covered.

Complete records of work performed are maintained by Transcat and are available for inspection. Laboratory standards used in the performance of this calibration are listed on this certificate.

Transcat documents the traceability of measurements to the SI units through the National Institute of Standards and Technology (NIST), or the National Research Council of Canada (NRC), or other national measurement institutes (NMI) that are signatories to the CIPM Mutual Recognition Arrangement, or accepted fundamental and/or natural physical constants, or by the use of specified methods, consensus standards or ratio type measurements. Documentation supporting traceability information is available for review upon written request at a Transcat facility. The measured quantity and the measurement uncertainty are required for further dissemination of traceability.

Uncertainties are reported with a coverage factor $k=2$, providing a level of confidence of approximately 95%. All calibrations have been performed using processes having a TUR of 4:1 or better (3:1 for mass calibrations), unless otherwise noted. The Test Uncertainty Ratio (TUR) is calculated in accordance with NCSL International RP-1B. For mass calibrations: Conventional mass referenced to 8.0 g/cm³.

The results in this report relate only to the item calibrated or tested. Recorded calibration data is valid at the time of calibration within the stated uncertainties at the environmental conditions noted. The determination of compliance to the specification is specific to the model/serial no./ID no. referenced above based on the tolerances shown; these tolerances are either the original equipment manufacturer's (OEM's) warranted specifications or the client's requested specifications. Any number of factors can cause a unit to drift out of tolerance at any time following its calibration. Limitations on the uses of this instrument are detailed in the OEM's operating instructions. This certificate may not be reproduced except in full, without the written approval of Transcat. Additional information, if applicable may be included on separate report(s).

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: 4302958578



Certificate/ISO Number: 5-E3K4K-20-1 Revision 0

As Found/As Left Data

| Description | Setpoints | Accuracy | Low Limit | High Limit | As Found / As Left | Cal Process | | Measurement Uncertainty (k=2; ±) | Units | TUR |
|---|-------------|--------------|-----------|------------|--------------------|-------------|---|----------------------------------|-------|----------|
| | | | | | | O | T | | | |
| Pressure Measure: 552 to 1172 mbara Range | | | | | | | | | | |
| | 550.53mbara | ±(0.015% FS) | 550.35 | 550.71 | 550.50 mbara | 1.4e-002 | | 1.5e-002 | mbara | 13.1 : 1 |
| | 610.62mbara | ±(0.015% FS) | 610.44 | 610.80 | 610.60 mbara | 1.5e-002 | | 1.6e-002 | mbara | 11.8 : 1 |
| | 680.95mbara | ±(0.015% FS) | 680.77 | 681.13 | 680.90 mbara | 1.7e-002 | | 1.8e-002 | mbara | 10.6 : 1 |
| | 742.77mbara | ±(0.015% FS) | 742.59 | 742.95 | 742.70 mbara | 1.9e-002 | | 1.9e-002 | mbara | 9.7 : 1 |
| | 803.04mbara | ±(0.015% FS) | 802.86 | 803.22 | 803.00 mbara | 2.0e-002 | | 2.1e-002 | mbara | 9.0 : 1 |
| | 863.43mbara | ±(0.015% FS) | 863.25 | 863.61 | 863.40 mbara | 2.2e-002 | | 2.2e-002 | mbara | 8.3 : 1 |
| | 923.57mbara | ±(0.015% FS) | 923.39 | 923.75 | 923.60 mbara | 2.3e-002 | | 2.4e-002 | mbara | 7.8 : 1 |
| | 983.80mbara | ±(0.015% FS) | 983.62 | 983.98 | 983.80 mbara | 2.5e-002 | | 2.6e-002 | mbara | 7.3 : 1 |
| | 1052.8mbara | ±(0.015% FS) | 1052.6 | 1053.0 | 1052.8 mbara | 2.6e-002 | | 6.3e-002 | mbara | 7.6 : 1 |
| | 1113.2mbara | ±(0.015% FS) | 1113.0 | 1113.4 | 1113.2 mbara | 2.8e-002 | | 6.5e-002 | mbara | 7.2 : 1 |
| | 1173.4mbara | ±(0.015% FS) | 1173.2 | 1173.6 | 1173.4 mbara | 2.9e-002 | | 6.5e-002 | mbara | 6.8 : 1 |
| | 923.57mbara | ±(0.015% FS) | 923.39 | 923.75 | 923.60 mbara | 2.3e-002 | | 2.4e-002 | mbara | 7.8 : 1 |
| | 863.43mbara | ±(0.015% FS) | 863.25 | 863.61 | 863.40 mbara | 2.2e-002 | | 2.2e-002 | mbara | 8.3 : 1 |
| | 803.04mbara | ±(0.015% FS) | 802.86 | 803.22 | 803.00 mbara | 2.0e-002 | | 2.1e-002 | mbara | 9.0 : 1 |

Field not applicable.

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: 4302958578



Certificate/SO Number: 5-E3K4K-20-1 Revision 0

Traceable Standards

| Asset | Manufacturer | Model Number | Description | Cal Date | Due Date | Traceability Number | Use |
|----------|----------------------|--------------------|----------------------------|-----------|-----------|---------------------|-------|
| DewK1 | Hart Scientific | 2626-H | Hygro-Thermometer, Probe, | 29-Jun-22 | 30-Jun-23 | 1-&DEWK1-12-1 | AF/AL |
| DW09BA | Fluke/DH Instruments | PG7601 | Piston Gauge | 8-Aug-22 | 31-Aug-23 | 5-&DW09BA-13-1 | AF/AL |
| DW09LOW | Fluke/DH Instruments | PC-7100/7600-10-TC | Gas Piston-Cylinder Module | 3-Jul-18 | 3-Jul-23 | 5-&DW09LOW-1-1 | AF/AL |
| DW09MASS | Fluke/DH Instruments | MS-AMH-38 | AMH Mass Set | 4-Jan-23 | 31-Jan-24 | 5-&DW09MASS-7-1 | AF/AL |

The use of the standard is defined as: AF - used for as-found readings, AL - used for as-left readings.

Environmental Data

| Temperature | Relative Humidity | Temp / RH Asset | Lab Area | Lab Description |
|-------------------|-------------------|-----------------|----------|-----------------|
| 71.74°F / 22.08°C | 51.00% | DewK9 | B | GP Pressure |

Decision Rule

When compliance statements are present, they are reported without factoring in the effects of uncertainty and comply with the guidelines as follows: The acceptance zone is defined as: less than or equal to the high limit, and/or greater than or equal to the low limit. The rejection zones are defined as greater than the high limit and/or less than the low limit. Single measurement results in the acceptance zone are identified as in-tolerance. Single measurement results in the rejection zone are identified as out-of-tolerance (OOT). When all measurement results are in the acceptance zone for repeated measurements, for the same characteristic, the test is identified as in-tolerance. For repeated characteristic measurements, a single measurement result in the rejection zone, will cause the test to be identified as out-of-tolerance (OOT). Data rejection for cause, (outliers) is permitted after the "Determining and Verifying Out Of Tolerance (OOT) and/or Op Fail Readings" procedure outlined in this document has been completed and the anomalous reading cannot be repeated, and the anomalous reading does not represent the system under test. Statements of conformity are binary.

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: 4302958578



Certificate/SO Number: 5-E3K4K-20-1 Revision 0

Legend

| Topic | Description |
|-------------------------------|--|
| Accuracy | UUT specification that establishes expected tolerances and a time limit (calibration interval) over which the instrument is expected to hold these tolerances |
| As Found | Initial measurement results |
| As Left | Measurement results after adjustment and/or repair |
| Blank Data Field | Test is not applicable for the UUT |
| Cal Process Uncertainty (CPU) | The uncertainty of calibration process for the reported measurement result |
| Calibration Date | Indicates the date that the calibration was completed |
| Cover Factor (K) | A measure of uncertainty that defines an interval about the measurement result |
| Due Date | Indicates the end of the calibration cycle as requested by the customer |
| Issue Date | Indicates the date that the calibration has passed the Data Review Process and was signed by an authorized signatory or the date that a revision to the original certificate has been issued |
| Low / High Limits | Establishes UUT acceptable performance limits for the test measurement |
| Measurement Uncertainty | The dispersion of the values attributed to a measured quantity |
| OOA | Out of Acceptance (#) |
| OOT | Out of Tolerance (*) |
| Setpoints | Measurement target values |
| Traceability | Unbroken chain of comparisons relating an instrument's measurements to a known standard(s) |
| Traceability Number | Unique identifier(s) used to document traceability of calibration standards |
| TUR | Test Uncertainty Ratio, ratio of the tolerance or specification of the test measurement in relation to the uncertainty in measurement results |
| UUT | Unit Under test |

CALIBRATED
BY **TRANSCAT**

CERTIFICATE OF CALIBRATION

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: 4302958578



Certificate/SO Number: 5-E3K4K-20-1 Revision 0

Calibrated At:
16115 Park Row
Houston, TX 77084

Facility Responsible:
16115 Park Row
Houston, TX 77084
800-828-1470



Date Received: March 13, 2023
Service Level: R9

| Calibrated By: | Reviewed By: |
|---|---------------------------------|
| Fritz Cardona Calibration Technician | Josh Soileau for |
| Mar 22, 2023 12:49:35 -04:00 | Mar 22, 2023 14:10:54 -04:00 |

Customer Number: 1-659111-00C
OPS-F20-014R10 09/29/21 FP001R9 4/9/2021

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Sales order: 1111663404

Date: July 05, 2022

NJSP DEPT OF LAW AND PUBLIC SAFETY

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer

ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.

CALGAZ LOT#: 302-402448281

ETHANOL IN NITROGEN

Product Expiration: May 19, 2025

| COMPONENT | PPM | (BrAC) |
|---------------------------------|------------|-----------------|
| ETHANOL | 260.5PPM | (0.100) |
| NITROGEN | BAL | |
| AVERAGE ANALYTICAL VALUE | PPM | (BrAC) |
| ETHANOL | 263.2 | (0.101) |

| REFERENCE STANDARD | CYLINDER | CONCENTRATION PPM |
|-----------------------------|----------|-------------------|
| N.M.I. TRACEABLE STANDARDS* | ND38434 | 260.4 |

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Calibration test 283190, 283189, 283188, or 283192 dated 6th January 2022 applies

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: ND38434-20211028, A679, A650, ND38462-20211027, ND18363-20211104, ND50144-20201218

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: May 19, 2022

APPROVED BY: _____



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410) 228-6400 Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Sales order: 1111918174

Date: July 27, 2022

DRAEGER MEDICAL SYSTEMS INC.

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer

ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.

CALGAZ LOT#: 302-402488140

ETHANOL IN NITROGEN

Product Expiration: July 15, 2025

| COMPONENT | PPM | (BrAC) |
|--------------------------|----------|----------|
| ETHANOL | 104.2PPM | (0.040) |
| NITROGEN | BAL | |
| AVERAGE ANALYTICAL VALUE | PPM | (BrAC) |
| ETHANOL | 107.1 | (0.041) |

| REFERENCE STANDARD | CYLINDER | CONCENTRATION PPM |
|-----------------------------|----------|-------------------|
| N.M.I. TRACEABLE STANDARDS* | ND38434 | 260.4 |

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Calibration test 283192, dated 6th January 2022 or calibration test 292029, 292030 or 292031, dated 26th March 2022 applies

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: ND38434-20211028, A679, A650, ND38462-20211027, ND18363-20211104, ND50144-20201218

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: July 15, 2022

APPROVED BY: _____



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410) 228-6400 Fax: (410) 228-4251



CERTIFICATE OF ANALYSIS
EBS - ETHANOL BREATH STANDARD

DRAEGER INC HOUSTON
7256 SOUTH SAM HOUSTON W PKWY
HOUSTON, TX 77085

INVOICE#: 21121898
PO#: US44302345493
CUST. ITEM #: 4401020
DATE: Apr. 14, 2021

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer
ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.
CALGAZ LOT#: 1474760
ETHANOL IN NITROGEN

PRODUCT EXPIRATION: Apr. 03, 2024

Table with 3 columns: COMPONENT, PPM, (BrAC). Rows include ETHANOL (208.4 PPM, 0.080 BrAC), NITROGEN (BAL), AVERAGE ANALYTICAL VALUE (211.0 PPM, 0.081 BrAC).

Table with 3 columns: REFERENCE STANDARD, CYLINDER, CONCENTRATION PPM. Row: N.M.I. TRACEABLE STANDARDS*, ND50144, 260.6

* CERTIFICATION TRACEABLE TO National Metrology Institute Traceable Standards.

TRACEABILITY

Preparation: Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44. Calibration test 121088, 121097, 121091, or 121100 dated, 18th January 2019 applies.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards. Certification Numbers: ND50144-20201218, A679, ND18363-20191203, A650

No affecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA). CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

MANUFACTURED DATE: Apr. 03, 2021

CALGAZ CYLINDER SIZE: 6DM

APPROVED BY : [Signature]

"We certify that all the cylinders for the lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Sales order: 1111788955
Date: July 14, 2022

NJSP

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer
 ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.
 CALGAZ LOT#: 302-402486003
 ETHANOL IN NITROGEN

Product Expiration: July 12, 2025

| COMPONENT | PPM | (BrAC) |
|---------------------------------|------------|-----------------|
| ETHANOL | 416.8PPM | (0.160) |
| NITROGEN | BAL | |
| AVERAGE ANALYTICAL VALUE | PPM | (BrAC) |
| ETHANOL | 418.2 | (0.161) |

| REFERENCE STANDARD | CYLINDER | CONCENTRATION PPM |
|-----------------------------|----------|-------------------|
| N.M.I. TRACEABLE STANDARDS* | ND38434 | 260.4 |

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Calibration test 283192, dated 6th January 2022 or calibration test 292029, 292030 or 292031, dated 26th March 2022 applies

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: ND38434-20211028, A679, A650, ND38462-20211027, ND18363-20211104, ND50144-20201218


No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: July 12, 2022

APPROVED BY: _____



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 170.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
 821 Chesapeake Drive, Cambridge, MD 21613-0149
 Phone: (410) 228-6400 Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS
EBS - ETHANOL BREATH STANDARD

DRAEGER INC HOUSTON
HOUSTON, TX 77085

REF#: 21775602
DOC#: US44302405855
CUST. ITEM #: 4401041
DATE: Jun. 23, 2021

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer
ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.
CALGAZ LOT#: 1495468
ETHANOL IN NITROGEN

PRODUCT EXPIRATION: May. 14, 2024

| COMPONENT | PPM | (BrAC) |
|-----------|-------|----------|
| ETHANOL | 781.5 | (0.300) |
| NITROGEN | BAL | |

| REFERENCE STANDARD | CYLINDER | CONCENTRATION PPM |
|-----------------------------|----------|-------------------|
| N.M.I. TRACEABLE STANDARDS* | ND50144 | 260.6 |

* Certification traceability is recognized by NIST through the CIPM MRA.

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.
Calibration test 219908, 219911, 219909, or 219926 dated, 6th January 2021 applies.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.
Certification Numbers: ND50144-20201218, A679, ND18363-20191203, A650

No affecting environmental conditions during analysis.


*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

MANUFACTURED DATE: May. 14, 2021

CALGAZ CYLINDER SIZE: 6D

APPROVED BY :



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410)228-6400 Fax: (410)228-4251

DEPARTMENT OF
Law and Public Safety
 This is to certify that

Dennis J. Lutz

New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 9510
 A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 8th DAY OF June

TWO THOUSAND AND Twenty One

[Signature]
 COLONEL
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

| | DATE | Refresher Course PLACE | INSTRUCTOR |
|----|---------|---------------------------|------------|
| 1. | 7/13/23 | Hamilton Tech | R. G. |
| 2. | | | |
| 3. | | | |
| 4. | | | |
| 5. | | | |
| 6. | | | |
| 7. | | | |
| 8. | | | |
| 9. | | | |

S.P. 293B (Rev. 01/18)

DEPARTMENT OF
Law and Public Safety
 This is to certify that

Dennis J. Lutz

Breath Test Coordinator/Instructor

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 9510
 A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 8th DAY OF June

TWO THOUSAND AND Twenty One

[Signature]
 COLONEL
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

| | DATE | Refresher Course PLACE | INSTRUCTOR |
|----|------|---------------------------|------------|
| 1. | | | |
| 2. | | | |
| 3. | | | |
| 4. | | | |
| 5. | | | |
| 6. | | | |
| 7. | | | |
| 8. | | | |
| 9. | | | |

S.P. 293B (Rev. 01/18)