

ALCOTEST CERTIFICATIONS INDEX # 06

01-02-2025

ALCOTEST 9510

SERIAL # ARMJ-0142

- ITEM # 1 9510 PARAMETER REPORT
- ITEM # 2 9510 CERT REPORT – WET ADJUST (PART I)
- ITEM # 3 9510 CERT REPORT – DRY ADJUST (PART II)
- ITEM # 4 9510 CERT REPORT – LINEARITY (PART III)
- ITEM # 5 9510 CYLINDER INSTALLATION REPORT – INLET 1
- ITEM # 6 9510 CYLINDER INSTALLATION REPORT – INLET 2
- ITEM # 7 CERT OF ANALYSIS, CALGAS LOT# 302-402921459
- ITEM # 8 CERT OF ANALYSIS, CALGAS LOT# 302-402843436
- ITEM # 9 CERT OF ACCURACY 9510, SERIAL # ARMJ-0142
- ITEM # 10 CERT OF ANALYSIS SIMULATOR SOLUTION LOT # 23240
- ITEM # 11 CERT OF CALIBRATION, CERT/SO#5-E8A6B-20-1 Revision 0 (5 pgs)
- ITEM # 12 CERT OF CALIBRATION, CERT/SO#5-E8A6B-280-1 Revision 0 (5 pgs)
- ITEM # 13 CERT OF ANALYSIS, CALGAS LOT# 302-402448282
- ITEM # 14 CERT OF ANALYSIS, CALGAS LOT# 302-402755169
- ITEM # 15 CERT OF ANALYSIS, CALGAS LOT# 302-402477284
- ITEM # 16 CERT OF ANALYSIS, CALGAS LOT# 302-402486005
- ITEM # 17 CERT OF ANALYSIS, CALGAS LOT# 302-402759888
- ITEM # 18 NJSP CERTIFICATION OFFICER CREDENTIALS (1 pgs)

ALCOTEST 9510 PARAMETER REPORT

Equipment

Serial No.: ARMJ-0142
Firmware: 8326739 1.5
WinCE application: 8326738 2.9
Configuration: 8326737 3.10

Date: 01/02/2025
Time: 09:51:45

Parameter

| | | |
|---|--------|-------------|
| min. blow time | 5.0 | s |
| min. breath volume for females of age 60+ | 1.2 | L |
| min. breath volume for all other | 1.5 | L |
| min. blow flow | 4.5 | L/min |
| plateau detection limit | 4 | % |
| plateau detection start conc. | 70 | microgram/L |
| neg. flow detection (part. vacuum) | 10.0 | hPa |
| neg. flow detection sensitivity | 10 | |
| cal. gas abort volume | 0.4 | L |
| result-to-zero limit | 0.0050 | %BAC |
| ambient air check limit | 0.0049 | %BAC |
| interference det. d-criterion limit abs. | 38 | microgram/L |
| interference det. d-criterion limit rel. | 10.0 | % |
| interference det. t-criterion limit abs. | 8 | microgram/L |
| interference det. t-criterion limit rel. | 2.1 | % |
| IR CO2 offset | 10 | microgram/L |
| IR H2O offset | 4 | microgram/L |
| EC H2O offset | 0 | microgram/L |
| Value-based EC aging comp. on/off (1/0) | 0 | |
| Time-based EC aging comp. on/off (1/0) | 1 | |
| Time-based EC aging comp. per month | 0.2 | % |
| Time-based EC aging comp. maximum | 3.0 | % |
| EC fatigue comp. max. sum | 15000 | |
| EC fatigue comp. factor | 50 | |
| EC fatigue comp. minutes | 180 | |
| mouth alc. mark limit | 500 | |
| mouth alc. lower limit | 30 | |
| mouth alc. slope | 6 | |
| mouth alc. zero limit | 50 | |
| mouth alc. max. neg. sum | 6 | |
| mouth alc. max. 2nd derivative | 35 | |

ALCOTEST 9510 CERTIFICATION REPORT - WET ADJUST (PART I)
Toms River Twp

Equipment

| | | | | | |
|------------------|---------------|-------------|--------------|--------|-------------|
| Inst. Model No.: | ALCOTEST 9510 | Serial No.: | ARMJ-0142 | | |
| Firmware: | 8326739 1.5 | Config.: | 8326737 3.10 | WinCE: | 8326738 2.9 |

Wet Adjust Record

| | | | | | |
|----------------------|-----|------------------|------------|-----------------|---|
| Wet Adjust File No.: | 397 | Wet Adjust Date: | 01/02/2025 | Wet Adjust No.: | 5 |
| | | Wet Adjust Time: | 10:49:53 | | |

| | | | | | |
|-------------------|------------|---------------------|-----------|--------------------|------------|
| Concentration: | 0.100 % | Adj. Unit Ser. No.: | ARND-0009 | Adj. Unit Exp.: | 02/16/2025 |
| Adjusting Unit: | X-Cal 2000 | Soln. Bottle No.: | 842 | Adjust Soln. Exp.: | 06/28/2025 |
| Solution Lot No.: | 23240 | | | | |

| | |
|-----------------------------|----------------|
| Preadjust Simulator Temp.: | 34.00 degree C |
| Postadjust Simulator Temp.: | 34.00 degree C |

Result

Procedure completed successfully.

Coordinator

| | | | |
|---------------------|-------------------|--------|-----------------|
| Last Name: Bellay - | First Name: David | MI: M. | Badge No.: 8112 |
|---------------------|-------------------|--------|-----------------|

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

Twp D J Bellay 8112

Signed:

Date: 01/02/2025

ID: 50

ALCOTEST 9510 CERTIFICATION REPORT - DRY ADJUST (PART II)
Toms River Twp

Equipment

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMJ-0142
Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9

Dry Adjust Record

Dry Adjust File No.: 398 Dry Adjust Date: 01/02/2025 Dry Adjust No.: 4
Dry Adjust Time: 11:15:17

Concentration: 0.100 %
Dry Gas Lot No.: 302-402448282 Adjust Gas Exp.: 05/20/2025
Barom. Model No.: Mensor CPG2300 Barom. Serial No.: 4100126V Barom. Cert. Exp.: 01/10/2025
Pre-adjust Amb. Pressure: 1013 hPa Post-adjust Amb. Pressure: 1013 hPa

Result

Procedure completed successfully.

Coordinator

Last Name: Bellay - First Name: David MI: M. Badge No.: 8112

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

Tp D J Bellay 8112

Signed:

Date: 01/02/2025

ID: 50

ALCOTEST 9510 CERTIFICATION REPORT - LINEARITY (PART III)
Toms River Twp

Equipment

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMJ-0142
Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9

Linearity Record

Linearity File No.: 399 Lin. Date: 01/02/2025 Lin. No.: 4

0.040% Dry Gas Lot No.: 302-402755169 Adjust. Gas Exp.: 05/25/2026
0.080% Dry Gas Lot No.: 302-402477284 Adjust. Gas Exp.: 06/27/2025
0.160% Dry Gas Lot No.: 302-402486005 Adjust. Gas Exp.: 07/13/2025
0.300% Dry Gas Lot No.: 302-402759888 Adjust. Gas Exp.: 05/31/2026

Data Summary

| Function | Result %BAC | Time hh:mm:ss | Barometric Pres. [hPa] | Comment(s) or Status Code |
|-----------------------|----------------|------------------|---------------------------|------------------------------|
| Ambient Air Blank | 0.000 | 11:35:45 | | *TEST PASSED* |
| Control .04 Test 1 EC | 0.039 | 11:36:20 | 1012 | *TEST PASSED* |
| Control .04 Test 1 IR | 0.039 | 11:36:20 | 1012 | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 11:37:27 | | *TEST PASSED* |
| Control .04 Test 2 EC | 0.039 | 11:37:38 | 1013 | *TEST PASSED* |
| Control .04 Test 2 IR | 0.040 | 11:37:38 | 1013 | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 11:39:08 | | *TEST PASSED* |
| Control .08 Test 3 EC | 0.078 | 11:39:41 | 1013 | *TEST PASSED* |
| Control .08 Test 3 IR | 0.079 | 11:39:41 | 1013 | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 11:40:52 | | *TEST PASSED* |
| Control .08 Test 4 EC | 0.080 | 11:41:04 | 1013 | *TEST PASSED* |
| Control .08 Test 4 IR | 0.080 | 11:41:04 | 1013 | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 11:42:36 | | *TEST PASSED* |
| Control .16 Test 5 EC | 0.156 | 11:43:10 | 1013 | *TEST PASSED* |
| Control .16 Test 5 IR | 0.158 | 11:43:10 | 1013 | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 11:44:28 | | *TEST PASSED* |
| Control .16 Test 6 EC | 0.159 | 11:44:40 | 1013 | *TEST PASSED* |
| Control .16 Test 6 IR | 0.160 | 11:44:40 | 1013 | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 11:51:05 | | *TEST PASSED* |
| Control .30 Test 7 EC | 0.300 | 11:51:40 | 1012 | *TEST PASSED* |
| Control .30 Test 7 IR | 0.304 | 11:51:40 | 1012 | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 11:53:07 | | *TEST PASSED* |
| Control .30 Test 8 EC | 0.307 | 11:53:19 | 1012 | *TEST PASSED* |
| Control .30 Test 8 IR | 0.307 | 11:53:19 | 1012 | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 11:53:55 | | *TEST PASSED* |

Result

All tests within acceptable tolerance.

Coordinator

Last Name: Bellay - First Name: David MI: M. Badge No.: 8112

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

Twp Bellay 8112

Signed:

Date: 01/02/2025

ID: 50

ALCOTEST 9510 CYLINDER INSTALLATION REPORT - INLET 1*Toms River Twp***SERIAL NUMBER: ARMJ-0142****Equipment**

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMJ-0142
Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9
Cyl1 Install File No.: 400 Cyl1 Install Date: 01/02/2025 Cyl1 Install No.: 4

Control Tests (0.100%)

Installation Inlet: #1 (Upper) Post test active Cyl.: #2 (Lower)
Dry Gas Lot No.: 302-402921459 Dry Gas Lot Exp.: 12/14/2026

Data Summary

| Function | Result %BAC | Time hh:mm:ss | Barometric Pres. [hPa] | Comment(s) or Status Code |
|-------------------|----------------|------------------|---------------------------|------------------------------|
| Ambient Air Blank | 0.000 | 12:12:13 | | *TEST PASSED* |
| Control Test 1 | | | 1012 | *TEST PASSED* |
| EC Result | 0.098 | 12:12:59 | | *TEST PASSED* |
| IR Result | 0.099 | 12:12:59 | | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 12:14:17 | | *TEST PASSED* |
| Control Test 2 | | | 1012 | *TEST PASSED* |
| EC Result | 0.099 | 12:14:41 | | *TEST PASSED* |
| IR Result | 0.099 | 12:14:41 | | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 12:15:58 | | *TEST PASSED* |
| Control Test 3 | | | 1012 | *TEST PASSED* |
| EC Result | 0.099 | 12:16:23 | | *TEST PASSED* |
| IR Result | 0.099 | 12:16:23 | | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 12:16:58 | | *TEST PASSED* |

Result

All tests within acceptable tolerance.

Coordinator

Last Name: Bellay - First Name: David MI: M. Badge No.: 8112

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

David Bellay 8112

Signed:

Date: 01/02/2025

ID: 50

ALCOTEST 9510 CYLINDER INSTALLATION REPORT - INLET 2
Tom's River Twp
SERIAL NUMBER: ARMJ-0142

Equipment

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMJ-0142
Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9
Cyl2 Install File No.: 256 Cyl2 Install Date: 07/02/2024 Cyl2 Install No.: 2

Control Tests (0.100%)

Installation Inlet: #2 (Lower) Post test active Cyl.: #1 (Upper)
Dry Gas Lot No.: 302-402843436 Dry Gas Lot Exp.: 09/08/2026

Data Summary

| Function | Result %BAC | Time hh:mm:ss | Barometric Pres. [hPa] | Comment(s) or Status Code |
|-------------------|----------------|------------------|---------------------------|------------------------------|
| Ambient Air Blank | 0.000 | 09:55:37 | | *TEST PASSED* |
| Control Test 1 | | | 1023 | *TEST PASSED* |
| EC Result | 0.101 | 09:56:23 | | *TEST PASSED* |
| IR Result | 0.099 | 09:56:23 | | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 09:57:33 | | *TEST PASSED* |
| Control Test 2 | | | 1023 | *TEST PASSED* |
| EC Result | 0.102 | 09:57:57 | | *TEST PASSED* |
| IR Result | 0.099 | 09:57:57 | | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 09:59:07 | | *TEST PASSED* |
| Control Test 3 | | | 1023 | *TEST PASSED* |
| EC Result | 0.102 | 09:59:31 | | *TEST PASSED* |
| IR Result | 0.100 | 09:59:31 | | *TEST PASSED* |
| Ambient Air Blank | 0.000 | 09:59:57 | | *TEST PASSED* |

Result

All tests within acceptable tolerance.

Coordinator

Last Name: Waldrop - First Name: Robert MI: W Badge No.: 8256

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

 #8256

Signed:

Date: 07/02/2024

ID: 52

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Part Number: 4401036

DRAEGER MEDICAL SYSTEMS INC

Sales order: 1126218824

Date: December 18, 2023

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer

ANALYTICAL ACCURACY: ± 0.002 BrAC or $\pm 2\%$ whichever is greater.

CALGAZ LOT#: 302-402921459

ETHANOL IN NITROGEN

Product Expiration: December 14, 2026

| COMPONENT | PPM | (BrAC) |
|--------------------------|----------|----------|
| ETHANOL | 260.5PPM | (0.100) |
| NITROGEN | BAL | |
| AVERAGE ANALYTICAL VALUE | PPM | (BrAC) |
| ETHANOL | 263.1 | (0.101) |

| REFERENCE STANDARD | CYLINDER | CONCENTRATION PPM |
|-----------------------------|----------|-------------------|
| N.M.I. TRACEABLE STANDARDS* | ND38424 | 260.7 |

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: December 14, 2023

APPROVED BY: 

"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC

821 Chesapeake Drive, Cambridge, MD 21613-0149

Phone: (410) 228-6400

Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Part Number: 4401036

DRAEGER MEDICAL SYSTEMS INC

Sales order: 1123816776

Date: September 18, 2023

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer

ANALYTICAL ACCURACY: ± 0.002 BrAC or $\pm 2\%$ whichever is greater.

CALGAZ LOT#: 302-402843436

ETHANOL IN NITROGEN

Product Expiration: September 08, 2026

| COMPONENT* | PPM | (BrAC) |
|--------------------------|----------|----------|
| ETHANOL | 260.5PPM | (0.100) |
| NITROGEN | BAL | |
| AVERAGE ANALYTICAL VALUE | PPM | (BrAC) |
| ETHANOL | 263.3 | (0.104) |

| REFERENCE STANDARD | CYLINDER | CONCENTRATION PPM |
|-----------------------------|----------|-------------------|
| N.M.I. TRACEABLE STANDARDS* | ND38424 | 260.7 |

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: September 08, 2023

APPROVED BY: 

"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 170.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC

821 Chesapeake Drive, Cambridge, MD 21613-0149

Phone: (410) 228-6400

Fax: (410) 228-4251



Dräger

Alcotest 9510

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 9510 has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest 9510 is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 49 FR 48864, and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your state's specifications.

Certification Date:

Serial Number:

11/12/2023

ARMJ-0142

Draeger, Inc.

AP MB



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

PHILIP D. MURPHY
Governor

TANESHA L. WAY
Lt. Governor

MATTHEW J. PLATKIN
Attorney General

COLONEL PATRICK J. CALLAHAN
Superintendent

CERTIFICATION OF ANALYSIS 0.100 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.

ANALYSIS DATE: 09/13/2023

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 23240

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1196 to 0.1212 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is June 28, 2025.

As OFS Director for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Michael Kennedy

Michael Kennedy
OFS Director
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 15 day of September, 2023.

Notary

KAREN E. STAHL
NOTARY PUBLIC OF NEW JERSEY
Commission # 50110522
My Commission Expires 6/13/2024



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable



Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085

PO Number: S1O4303208869

Certificate/SO Number: 5-E8A6B-120-1 Revision 0

Manufacturer: Drager Safety AG & Co. KGaA
Model Number: X-Cal 2000
Description: Breath Alcohol Simulator
Serial Number: ARND-0009
ID: NONE

As-Found: In Tolerance
As-Left: In Tolerance

Issue Date: Feb 16, 2024
Calibration Date: Feb 16, 2024
Due Date: Feb 16, 2025

Calibrated To: Customer Spec
Calibration Procedure: 1-AC103519-1

Transcat Calibration Laboratories have been audited and found in compliance with ISO /IEC 17025:2017. Accredited calibrations performed within the Lab Scope of Accreditation are indicated by the presence of the Transcat Logo and Certificate Number. Any measurements on an accredited calibration not covered by the Lab Scope of Accreditation are listed in the notes section of the certificate. SCC, NRC, CLAS or ANAB do not certify individual calibrations by accredited laboratories.

Transcat calibrations, as applicable, are performed in compliance with the requirements of the Transcat Quality Manual QAC-P01-000, the customer Purchase Order and/or Quality Agreement requirements Z540.1-1994 (R2002), and ISO 10012:2003, as applicable. When specified contractually, the requirements of ISO TS16949:2009, 10CFR21, 10CFR50 App. B, ASME NQA-1:2012, and ANSI/NCSL Z540.3 are covered.

Complete records of work performed are maintained by Transcat and are available for inspection. Laboratory standards used in the performance of this calibration are listed on this certificate.

Transcat documents the traceability of measurements to the SI units through the National Institute of Standards and Technology (NIST), or the National Research Council of Canada (NRC), or other nations (NMI) that are signatories to the CIPM Mutual Recognition Arrangement, or accepted fundamental and/or natural physical constants, or by the use of specified methods, consensus standards or ratio type methods. Documentation supporting traceability information is available for review upon written request at a Transcat facility. The measured quantity and the measurement uncertainty are required for further dissemination.

Uncertainties are reported with a coverage factor $k=2$, providing a level of confidence of approximately 95%. All calibrations have been performed using processes having a TUR of 4:1 or better (3:1 for mass calibrations) unless otherwise noted. The Test Uncertainty Ratio (TUR) is calculated in accordance with NCSL International RP-18. For mass calibrations: Conventional mass referenced to 8.0 g/cm³.

The results in this report relate only to the item calibrated or tested. Recorded calibration data is valid at the time of calibration within the stated uncertainties at the environmental conditions noted. The details of the specification are specific to the model/serial no./ID no. referenced above based on the tolerances shown; these tolerances are either the original equipment manufacturers (OEM's) warranted specifications or specifications. Any number of factors can cause a unit to drift out of tolerance at any time following its calibration. Limitations on the uses of this instrument are detailed in the OEM's operating instructions and should be reproduced except in full, without the written approval of Transcat. Additional information, if applicable may be included on separate report(s).

Date Received: January 04, 2024
Service Level: R9

Certificate - Page 1 of 5
Reprinted on February 27, 2024

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085

PO Number: S1O4303208869

Certificate/SO Number: 5-E8A6B-120-1 Revision 0

As Found/As Left Data

| Description | Setpoints | Accuracy | Low Limit | High Limit | As Found / As Left | O O T | Cal Proces Uncertain (k=2; ±) |
|---|-----------|-------------|-----------|------------|--------------------|-------------|-------------------------------------|
| Function Checks | | | | | | | |
| Bubble Check | | | P | P | P | | |
| Seal Check | | | P | P | P | | |
| Temperature Source: Accuracy Test | | | | | | | |
| Accuracy Test | 34.00°C | ±(0.02 °C) | 33.98 | 34.02 | 34.01 °C | | 1.5e-002 |
| Temperature Source: Stability Test | | | | | | | |
| Stability Test | 0.00°C | ±(0.02 °C) | -0.02 | 0.02 | 0.00 °C | | 5.0e-003 |

Traceable Standards

| Asset | Manufacturer | Model Number | Description | Cal Date | Due Date |
|----------|-----------------------|--------------|-------------------|-----------|-----------|
| 05H1431 | AccuMac Corporation | AM1760 | Secondary SPRT | 12-Feb-24 | 28-Feb-25 |
| HP927312 | Hart Scientific/Fluke | 1575 | Super Thermometer | 6-Dec-22 | 30-Jun-24 |

The use of the standard is defined as: AF - used for as-found readings, AL - used for as-left readings.

Environmental Data

| Temperature | Relative Humidity | Temp / RH Asset | Lab Area |
|------------------|-------------------|-----------------|----------|
| 70.09°F /21.16°C | 47.50% | Dewk15 | G |

Decision Rule

When compliance statements are present, they are reported without factoring in the effects of uncertainty and comply with the guidelines as follows : The acceptance to the high limit, and/or greater than or equal to the low limit. The rejection zones are defined as greater than the high limit and/or less than the low limit. Single measu

Date Received: January 04, 2024
Service Level : R9

Certificate - Page 2 of 5
Reprinted on February 27, 2024

Customer: DRAEGER INC

7256 S SAM HOUSTON PKWY W

STE 100

HOUSTON, TX 77085

PO Number: S1O4303208869

Certificate/SO Number: 5-E8A6B-120-1 Revision 0

are identified as in-tolerance. Single measurement results in the rejection zone are identified as out-of-tolerance (OOT). When all measurement results are in the measurements, for the same characteristic, the test is identified as in-tolerance. For repeated characteristic measurements, a single measurement result in the rejection zone is identified as out-of-tolerance (OOT). Data rejection for cause, (outliers) is permitted after the Determining and Verifying Out Of Tolerance (OOT) and/or Op Fail document has been completed and the anomalous reading cannot be repeated, and the anomalous reading does not represent the system under test. States

Date Received: January 04, 2024

Service Level : R9

Certificate - Page 3 of 5

Reprinted on February 27, 2024

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085

PO Number: S104303208869

Certificate/SO Number: 5-E8A6B-120-1 Revision 0

Legend

| Topic | Description |
|-------------------------------|---|
| Accuracy | UUT specification that establishes expected tolerances and a time limit (calibration interval) over which the instrument is expected to hold the |
| As Found | Initial measurement results |
| As Left | Measurement results after adjustment and/or repair |
| Blank Data Field | Test is not applicable for the UUT |
| Cal Process Uncertainty (CPU) | The uncertainty of calibration process for the reported measurement result |
| Calibration Date | Indicates the date that the calibration was completed |
| Cover Factor (k) | A measure of uncertainty that defines an interval about the measurement result |
| Due Date | Indicates the end of the calibration cycle as requested by the customer |
| Issue Date | Indicates the date that the calibration has passed the Data Review Process and was signed by an authorized signatory or the date that a rev has been issued |
| Low / High Limits | Establishes UUT acceptable performance limits for the test measurement |
| Measurement Uncertainty | The dispersion of the values attributed to a measured quantity |
| OOA | Out of Acceptance (#) |
| OOT | Out of Tolerance (*) |
| Setpoints | Measurement target values |
| Traceability | Unbroken chain of comparisons relating an instrument's measurements to a known standard(s) |
| Traceability Number | Unique identifier(s) used to document traceability of calibration standards |
| TUR | Test Uncertainty Ratio, ratio of the tolerance or specification of the test measurement in relation to the uncertainty in measurement results |
| UUT | Unit Under test |

Date Received: January 04, 2024
Service Level : R9

Certificate - Page 4 of 5
Reprinted on February 27, 2024


Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085

PO Number: S1O4303208869

Certificate/SO Number: 5-E8A6B-120-1 Revision 0

Calibrated At:
16115 Park Row
Houston, TX 77084

Facility Responsible:
16115 Park Row
Houston, TX 77084
800-828-1470

Calibrated By:
 **Electronically Signed By:**
Camden Alford

Unit Barcode: 
0900B541805

Camden Alford Feb 16, 2024
Calibration Technician 08:18:11 -05:00

Date Received: January 04, 2024
Service Level : R9

Certificate - Page 5 of 5
Reprinted on February 27, 2024

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085

PO Number: S1O4303208869

Certificate/SO Number: 5-E8A6B-220-1 Revision 0

Manufacturer: Wika Instr/Mensor Corp/Trend
Model Number: CPG2300
Description: Portable Barometer
Serial Number: 4100126V
ID: NONE

As-Found: In Tolerance
As-Left: In Tolerance

Issue Date: Jan 10, 2024
Calibration Date: Jan 10, 2024
Due Date: Jan 10, 2025

Calibrated To: Manufacturer S
Calibration Procedure: 1-AC107288-0

Transcat Calibration Laboratories have been audited and found in compliance with ISO /IEC 17025:2017. Accredited calibrations performed within the Lab Scope of Accreditation are indicated by the presence of the Transcat Logo and Certificate Number. Any measurements on an accredited calibration not covered by the Lab Scope of Accreditation are listed in the notes section of the certificate. SCC, NRC, CLAS or ANAB do not certify individual calibrations by accredited laboratories.

Transcat calibrations, as applicable, are performed in compliance with the requirements of the Transcat Quality Manual QAC-P01-000, the customer Purchase Order and/or Quality Agreement requirements Z540.1-1994 (R2002), and ISO 10012:2003, as applicable. When specified contractually, the requirements of ISO TS16949:2009, 10CFR21, 10CFR50 App. B, ASME NQA-1:2012, and ANSI/NCCL Z540.3 are covered.

Complete records of work performed are maintained by Transcat and are available for inspection. Laboratory standards used in the performance of this calibration are listed on this certificate.

Transcat documents the traceability of measurements to the SI units through the National Institute of Standards and Technology (NIST), or the National Research Council of Canada (NRC), or other national metrology institutes (NMI) that are signatories to the CIPM Mutual Recognition Arrangement, or accepted fundamental and/or natural physical constants, or by the use of specified methods, consensus standards or ratio type methods. Documentation supporting traceability information is available for review upon written request at a Transcat facility. The measured quantity and the measurement uncertainty are required for further dissemination.

Uncertainties are reported with a coverage factor $k=2$, providing a level of confidence of approximately 95%. All calibrations have been performed using processes having a TUR of 4:1 or better (3:1 for mass calibrations) unless otherwise noted. The Test Uncertainty Ratio (TUR) is calculated in accordance with NCCL International RP-18. For mass calibrations: Conventional mass referenced to 8.0 g/cm³.

The results in this report relate only to the item calibrated or tested. Recorded calibration data is valid at the time of calibration within the stated uncertainties at the environmental conditions noted. The data is specific to the model/serial no./ID no. referenced above based on the tolerances shown; these tolerances are either the original equipment manufacturers' (OEM's) warranted specifications or specifications. Any number of factors can cause a unit to drift out of tolerance at any time following its calibration. Limitations on the uses of this instrument are detailed in the OEM's operating instructions and should be reproduced except in full, without the written approval of Transcat. Additional information, if applicable, may be included on separate report(s).

Date Received: January 04, 2024
Service Level: R9

Certificate - Page 1 of 5
Reprinted on February 27, 2024

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085

PO Number: S1O4303208869

Certificate/SO Number: 5-E8A6B-220-1 Revision 0

As Found/As Left Data

| Description | Setpoints | Accuracy | Low Limit | High Limit | As Found / As Left | O O T | Cal Proces Uncertaint (k=2; ±) |
|---|-------------|--------------|-----------|------------|--------------------|-------------|--------------------------------------|
| Pressure Measure: 552 to1172 mbara Range | | | | | | | |
| | 550.56mbara | ±(0.015% FS) | 550.38 | 550.74 | 550.50 mbara | | 1.1e-002 |
| | 610.66mbara | ±(0.015% FS) | 610.48 | 610.84 | 610.60 mbara | | 1.2e-002 |
| | 670.94mbara | ±(0.015% FS) | 670.76 | 671.12 | 670.90 mbara | | 1.3e-002 |
| | 742.81mbara | ±(0.015% FS) | 742.63 | 742.99 | 742.70 mbara | | 1.5e-002 |
| | 803.09mbara | ±(0.015% FS) | 802.91 | 803.27 | 803.00 mbara | | 1.6e-002 |
| | 863.48mbara | ±(0.015% FS) | 863.30 | 863.66 | 863.40 mbara | | 1.7e-002 |
| | 923.61mbara | ±(0.015% FS) | 923.43 | 923.79 | 923.60 mbara | | 1.8e-002 |
| | 983.85mbara | ±(0.015% FS) | 983.67 | 984.03 | 983.80 mbara | | 2.0e-002 |
| | 1052.8mbara | ±(0.015% FS) | 1052.6 | 1053.0 | 1052.8 mbara | | 2.1e-002 |
| | 1113.2mbara | ±(0.015% FS) | 1113.0 | 1113.4 | 1113.2 mbara | | 2.2e-002 |
| | 1173.5mbara | ±(0.015% FS) | 1173.3 | 1173.7 | 1173.5 mbara | | 2.3e-002 |
| | 923.61mbara | ±(0.015% FS) | 923.43 | 923.79 | 923.60 mbara | | 1.8e-002 |
| | 863.48mbara | ±(0.015% FS) | 863.30 | 863.66 | 863.40 mbara | | 1.7e-002 |
| | 803.09mbara | ±(0.015% FS) | 802.91 | 803.27 | 803.00 mbara | | 1.6e-002 |

Date Received: January 04, 2024
Service Level : R9

Certificate - Page 2 of 5
Reprinted on February 27, 2024

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085

PO Number: S1O4303208869

Certificate/SO Number: 5-E8A6B-220-1 Revision 0

Traceable Standards

| Asset | Manufacturer | Model Number | Description | Cal Date | Due Date |
|----------|----------------------|--------------------|----------------------------|-----------|-----------|
| DewK2 | Hart Scientific | 2626-H | Hygro-Thermometer, Probe, | 8-Mar-23 | 31-Mar-24 |
| DW09BA | Fluke/DH Instruments | PG7601 | Piston Gauge | 11-Sep-23 | 30-Sep-24 |
| DW09LOW | Fluke/DH Instruments | PC-7100/7600-10-TC | Gas Piston-Cylinder Module | 22-Aug-23 | 31-Aug-28 |
| DW09MASS | Fluke/DH Instruments | MS-AMH-38 | AMH Mass Set | 4-Jan-23 | 31-Jan-24 |

The use of the standard is defined as: AF - used for as-found readings, AL - used for as-left readings.

Environmental Data

| Temperature | Relative Humidity | Temp / RH Asset | Lab Area |
|------------------|-------------------|-----------------|----------|
| 70.06°F /21.14°C | 29.40% | DewK8 | B |

Decision Rule

When compliance statements are present, they are reported without factoring in the effects of uncertainty and comply with the guidelines as follows : The acceptance to the high limit, and/or greater than or equal to the low limit. The rejection zones are defined as greater than the high limit and/or less than the low limit. Single measurements are identified as in-tolerance. Single measurement results in the rejection zone are identified as out-of-tolerance (OOT). When all measurement results are in the measurements, for the same characteristic, the test is identified as in-tolerance. For repeated characteristic measurements, a single measurement result in the rejection zone is identified as out-of-tolerance (OOT). Data rejection for cause, (outliers) is permitted after the Determining and Verifying Out Of Tolerance (OOT) and/or Op Fail document has been completed and the anomalous reading cannot be repeated, and the anomalous reading does not represent the system under test. Stater

Date Received: January 04, 2024
Service Level : R9

Certificate - Page 3 of 5

Reprinted on February 27, 2024

Customer: DRAEGER INC

7256 S SAM HOUSTON PKWY W

STE 100

HOUSTON, TX 77085

PO Number: S1O4303208869

Certificate/SO Number: 5-E8A6B-220-1 Revision 0

Legend

| Topic | Description |
|-------------------------------|---|
| Accuracy | UUT specification that establishes expected tolerances and a time limit (calibration interval) over which the instrument is expected to hold the |
| As Found | Initial measurement results |
| As Left | Measurement results after adjustment and/or repair |
| Blank Data Field | Test is not applicable for the UUT |
| Cal Process Uncertainty (CPU) | The uncertainty of calibration process for the reported measurement result |
| Calibration Date | Indicates the date that the calibration was completed |
| Cover Factor (k) | A measure of uncertainty that defines an interval about the measurement result |
| Due Date | Indicates the end of the calibration cycle as requested by the customer |
| Issue Date | Indicates the date that the calibration has passed the Data Review Process and was signed by an authorized signatory or the date that a rev has been issued |
| Low / High Limits | Establishes UUT acceptable performance limits for the test measurement |
| Measurement Uncertainty | The dispersion of the values attributed to a measured quantity |
| OOA | Out of Acceptance (#) |
| OOT | Out of Tolerance (*) |
| Setpoints | Measurement target values |
| Traceability | Unbroken chain of comparisons relating an instrument's measurements to a known standard(s) |
| Traceability Number | Unique identifier(s) used to document traceability of calibration standards |
| TUR | Test Uncertainty Ratio, ratio of the tolerance or specification of the test measurement in relation to the uncertainty in measurement results |
| UUT | Unit Under test |

Date Received: January 04, 2024

Service Level: R9

Certificate - Page 4 of 5

Reprinted on February 27, 2024

Customer: DRAEGER INC

7256 S SAM HOUSTON PKWY W

STE 100


HOUSTON, TX 77085

PO Number: S1O4303208869

Certificate/SO Number: 5-E8A6B-220-1 Revision 0

Calibrated At:
16115 Park Row
Houston, TX 77084

Facility Responsible:
16115 Park Row
Houston, TX 77084
800-828-1470

Calibrated By:
 **Electronically Signed By:**
Fritz Cardona

Fritz Cardona Jan 10, 2024
Calibration Technician 13:09:11 -05:00

Unit Barcode:



0900B541818

Date Received: January 04, 2024
Service Level : R9

Certificate - Page 5 of 5

Reprinted on February 27, 2024

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Sales order: 1111663404
Date: July 05, 2022

NJSP DEPT OF LAW AND PUBLIC SAFETY

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer

ANALYTICAL ACCURACY: ± 0.002 BrAC or $\pm 2\%$ whichever is greater.

CALGAZ LOT#: 302-402448282

ETHANOL IN NITROGEN

Product Expiration: May 20, 2025

| COMPONENT | PPM | (BrAC) |
|--------------------------|----------|----------|
| ETHANOL | 260.5PPM | (0.100) |
| NITROGEN | BAL | |
| AVERAGE ANALYTICAL VALUE | PPM | (BrAC) |
| ETHANOL | 263.3 | (0.101) |

| REFERENCE STANDARD | CYLINDER | CONCENTRATION PPM |
|-----------------------------|----------|-------------------|
| N.M.I. TRACEABLE STANDARDS* | ND38434 | 260.4 |

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Calibration test 283190, 283189, 283188, or 283192 dated 6th January 2022 applies

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: ND38434-20211028, A679, A650, ND38462-20211027, ND18363-20211104, ND50144-20201218


No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: May 20, 2022

APPROVED BY:



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410) 228-6400 Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

DRAEGER MEDICAL SYSTEMS INC.;

Sales order: 1121156486
Date: June 12, 2023

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer

ANALYTICAL ACCURACY: ± 0.002 BrAC or $\pm 2\%$ whichever is greater.

CALGAZ LOT#: 302-402755169

ETHANOL IN NITROGEN

Product Expiration: May 25, 2026

| COMPONENT | PPM | (BrAC) |
|--------------------------|----------|----------|
| ETHANOL | 104.2PPM | (0.040) |
| NITROGEN | BAL | |
| AVERAGE ANALYTICAL VALUE | PPM | (BrAC) |
| ETHANOL | 107.2 | (0.041) |

| REFERENCE STANDARD | CYLINDER | CONCENTRATION PPM |
|-----------------------------|----------|-------------------|
| N.M.I. TRACEABLE STANDARDS* | ND38424 | 260.7 |

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: May 25, 2023

APPROVED BY: 

"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 170.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410) 228-6400 Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

DRAEGER MEDICAL SYSTEMS INC.;

Sales order: 1113360565

Date: August 29, 2022

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer

ANALYTICAL ACCURACY: ± 0.002 BrAC or $\pm 2\%$ whichever is greater.

CALGAZ LOT#: 302-402477284

ETHANOL IN NITROGEN

Product Expiration: June 27, 2025

| COMPONENT | PPM | (BrAC) |
|--------------------------|----------|----------|
| ETHANOL | 208.4PPM | (0.080) |
| NITROGEN | BAL | |
| AVERAGE ANALYTICAL VALUE | PPM | (BrAC) |
| ETHANOL | 210.9 | (0.081) |

| REFERENCE STANDARD | CYLINDER | CONCENTRATION PPM |
|-----------------------------|----------|-------------------|
| N.M.I. TRACEABLE STANDARDS* | ND38434 | 260.4 |

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Calibration test 283192, dated 6th January 2022 or calibration test 292029, 292030 or 292031, dated 26th March 2022 applies

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: ND38434-20211028, A679, A650, ND38462-20211027, ND18363-20211104, ND50144-20201218

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: June 27, 2022

APPROVED BY: 

"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 170.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC

821 Chesapeake Drive, Cambridge, MD 21613-0149

Phone: (410) 228-6400

Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

NJSP

Sales order: 1111788955

Date: July 14, 2022

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer

ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.

CALGAZ LOT#: 302-402486005

ETHANOL IN NITROGEN

Product Expiration: July 13, 2025

| COMPONENT | PPM | (BrAC) |
|--------------------------|----------|----------|
| ETHANOL | 416.8PPM | (0.160) |
| NITROGEN | BAL | |
| AVERAGE ANALYTICAL VALUE | PPM | (BrAC) |
| ETHANOL | 420.0 | (0.161) |

| REFERENCE STANDARD | CYLINDER | CONCENTRATION PPM |
|-----------------------------|----------|-------------------|
| N.M.I. TRACEABLE STANDARDS* | ND38434 | 260.4 |

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Calibration test 283192, dated 6th January 2022 or calibration test 292029, 292030 or 292031, dated 26th March 2022 applies

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: ND38434-20211028, A679, A650, ND38462-20211027, ND18363-20211104, ND50144-20201218


No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: July 13, 2022

APPROVED BY: _____



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC

821 Chesapeake Drive, Cambridge, MD 21613-0149

Phone: (410) 228-6400

Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Part Number: 4401041NJ
DRAEGER MEDICAL SYSTEMS INC

Sales order: 1123022087
Date: August 18, 2023

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer
ANALYTICAL ACCURACY: ± 0.002 BrAC or $\pm 2\%$ whichever is greater.
CALGAZ LOT#: 302-402759888
ETHANOL IN NITROGEN

Product Expiration: May 31, 2026

| COMPONENT | PPM | (BrAC) |
|--------------------------|----------|----------|
| ETHANOL | 781.5PPM | (0.300) |
| NITROGEN | BAL | |
| AVERAGE ANALYTICAL VALUE | PPM | (BrAC) |
| ETHANOL | 793.1 | (0.304) |

| REFERENCE STANDARD | CYLINDER | CONCENTRATION PPM |
|-----------------------------|----------|-------------------|
| N.M.I. TRACEABLE STANDARDS* | ND38424 | 260.7 |

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: A679-20190918, D049803-20220329


No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.


Manufactured Date: May 31, 2023

APPROVED BY:



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410) 228-6400 Fax: (410) 228-4251



David M. Bellay

New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF
THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 9510
A METHOD TO DETERMINE INTOXICATION.
GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 28th DAY OF April

TWO THOUSAND AND Twenty Three


David M. Bellay
COLONEL
NEW JERSEY STATE POLICE

Mr. J. A.
ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES

| DATE | Refresher Course PLACE | INSTRUCTOR |
|------|---------------------------|------------|
| 1. | | |
| 2. | | |
| 3. | | |
| 4. | | |
| 5. | | |
| 6. | | |
| 7. | | |
| 8. | | |
| 9. | | |

S.P. 293B (Rev. 10/22)



David M. Bellay

Breath Test Coordinator/Instructor

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF
THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 9510
A METHOD TO DETERMINE INTOXICATION.
GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 20th DAY OF August

TWO THOUSAND AND Twenty Four

David M. Bellay
COLONEL
NEW JERSEY STATE POLICE

Mr. J. A.
ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES

| DATE | Refresher Course PLACE | INSTRUCTOR |
|------|---------------------------|------------|
| 1. | | |
| 2. | | |
| 3. | | |
| 4. | | |
| 5. | | |
| 6. | | |
| 7. | | |
| 8. | | |
| 9. | | |

S.P. 293B (Rev. 10/22)

DEPARTMENT OF
Motor and Public Safety
This is to certify that

Robert W. Waldrop

Breath Test Coordinator/Instructor

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1966 IN THE OPERATION OF THE **Alcotest 9510**

A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 25th DAY OF March

TWO THOUSAND AND Twenty Four

[Signature]
COLONEL
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES

| DATE | Refresher Course PLACE | INSTRUCTOR |
|------|---------------------------|------------|
| 1. | | |
| 2. | | |
| 3. | | |
| 4. | | |
| 5. | | |
| 6. | | |
| 7. | | |
| 8. | | |
| 9. | | |

S.P. 293B (Rev. 10/22)

DEPARTMENT OF
Motor and Public Safety
This is to certify that

Robert W. Waldrop

New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1966 IN THE OPERATION OF THE **Alcotest 9510**

A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 28th DAY OF April

TWO THOUSAND AND Twenty Three

[Signature]
COLONEL
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES

| DATE | Refresher Course PLACE | INSTRUCTOR |
|------|---------------------------|------------|
| 1. | | |
| 2. | | |
| 3. | | |
| 4. | | |
| 5. | | |
| 6. | | |
| 7. | | |
| 8. | | |
| 9. | | |

S.P. 293B (Rev. 10/22)